



UITDRUKKING VAN DE MEETONZEKERHEID IN KALIBRATIE

De versies van documenten van het managementsysteem van BELAC die beschikbaar zijn op de website van BELAC (www.belac.be) worden beschouwd als de enige geldige versies.

Van toepassing vanaf: 07.06.2022

HISTORIEK VAN HET DOCUMENT

Herziening en datum van goedkeuring	Reden van de herziening	Draagwijdte van de herziening
EA-4/02 December 1999	Toepassing van het document EA-4/02. De betrokken instellingen en auditoren/experten worden verzocht om kennis te nemen van dit document op de website van EA.	
BELAC 2-107 Rev 0 CC 19.12.2006	Het document EA-4/02 wordt in zijn oorspronkelijke Engelstalige versie formeel opgenomen in de BELAC-documenten als document BELAC 2-107.	
BELAC 2-107 Rev 1 Secretariaat 23.06.2014	De gereviseerde versie van het document EA-4/02 rev1 wordt integraal overgenomen in dit document ter vervanging van EA-4/02 rev0.	
BELAC 2-107 Rev 2 Secretariaat 23.12.2021	De gereviseerde versie van het document EA-4/02 rev2 wordt integraal overgenomen in dit document ter vervanging van EA-4/02 rev1.	Geen significante wijzigingen: kleine verduidelijkingen en verbeteringen, correctie van kleine fouten, verwijzingen naar ISO/IEC 17025:2017 en toevoeging van een verwijzing naar beslisregels
BELAC 2-107 Rev 3 Secretariaat 07.06.2022	De gereviseerde versie van het document EA-4/02 rev3 wordt integraal overgenomen in dit document ter vervanging van EA-4/02 rev2.	Correctie redactionele fout op pagina 25 F4 EA-4/02 rev 02

INHOUDSOPGAVE

1	DOEL EN NORMATIEVE VERWIJZINGEN	4
2	BESTEMMELINGEN	4

UITDRUKKING VAN DE MEETONZEKERHEID IN KALIBRATIE

Nota:

BELAC past de documenten toe die EA, ILAC en IAF hebben uitgegeven met als doel de werking van de accreditatie-instellingen en de eisen gesteld aan de accreditatie-instellingen te harmoniseren.

De documenten die door EA, ILAC en IAF als verplicht worden opgelegd voor de ondertekening van de wederzijdse erkenningen zijn formeel geïntegreerd in de BELAC documenten

- *hetzij door de inhoud te verwerken;*
- *hetzij door, met akkoord van de Coördinatiecommissie, de volledige tekst in zijn originele versie over te nemen.*

1 DOEL EN NORMATIEVE VERWIJZINGEN

De norm EN ISO/IEC 17025 (§ 7.6.2) eist van een kalibratielaboratorium en eveneens van een testlaboratorium dat zijn eigen kalibraties uitvoert dat ze over een procedure beschikken om de meetonzekerheid van alle kalibraties en alle types van kalibratie te schatten.

Dit document preciseert de eisen van de norm voor wat betreft de uitdrukking van de meetonzekerheid in kalibratie.

Het herneemt de volledige tekst van document EA-4/02: Rev3 2022.

2 BESTEMMELINGEN

Met opvolging van de actualiseringen:

- De leden van de Coördinatiecommissie
- De leden van het Accreditatiebureau
- Het Accreditatiesecretariaat
- De auditoren voor de sector laboratoria
- De geaccrediteerde kalibratie- en testlaboratoria

Zonder opvolging van de actualiseringen:

- Elke aanvrager

**Publication
Reference**

EA-4/02 M: 2022

Evaluation of the Uncertainty of Measurement in calibration

PURPOSE

The purpose of this document is to harmonise evaluation of uncertainty of measurement within EA, to set up, in addition to the general requirements of EA, the specific demands in reporting uncertainty of measurement on calibration certificates issued by accredited laboratories and to assist accreditation bodies with a coherent assignment of Calibration and Measurement Capability to calibration laboratories accredited by them. As the rules laid down in this document are in compliance with both ILAC policy for uncertainty in calibration and the recommendations of the Guide to the Expression of Uncertainty in Measurement, the implementation of EA-4/02 will also foster the global acceptance of European results of measurement.

Authorship

This document has been drafted by the EA Laboratory Committee. This edition of EA-4/02 incorporates minor clarifications and enhancements, correction of minor errors, references to ISO/IEC 17025:2017 and inclusion of reference to Decision Rules.

Official language

The text may be translated into other languages as required. The English language version remains the definitive version.

Copyright

The copyright of this text is held by EA. The text may not be copied for resale.

Further information

For further information about this publication, please contact the Secretariat.

Please check our website for up-to-date information <http://www.european-accreditation.org>

Category: Application document with mandatory status

Date of approval: 5th November 2021

(Note: EA-4/02 was re-published (4th April 2022_rev 03) following an editorial mistake on page 25 F4 in the version (15th November 2021_rev 02) approved 5th November 2021.

Date of implementation: Immediate

Transitional period: None

CONTENTS

1	INTRODUCTION	4
2	OUTLINE AND DEFINITIONS	4
3	EVALUATION OF THE UNCERTAINTY OF MEASUREMENT OF INPUT ESTIMATES	6
4	CALCULATION OF THE STANDARD UNCERTAINTY OF THE OUTPUT ESTIMATE	9
5	EXPANDED UNCERTAINTY OF MEASUREMENT	12
6	STEP-BY-STEP PROCEDURE FOR CALCULATING THE UNCERTAINTY OF MEASUREMENT	13
7	REFERENCES	14
	APPENDIX A: CALIBRATION AND MEASUREMENT CAPABILITY	15
	APPENDIX B: GLOSSARY OF SOME RELEVANT TERMS	16
	APPENDIX C: SOURCES OF UNCERTAINTY OF MEASUREMENT	19
	APPENDIX D: CORRELATED INPUT QUANTITIES	20
	APPENDIX E: COVERAGE FACTORS DERIVED FROM EFFECTIVE DEGREES OF FREEDOM	23
	APPENDIX F: CONFORMITY WITH A SPECIFICATION	25
	SUPPLEMENT 1	27
S1	INTRODUCTION	28
S2	CALIBRATION OF A WEIGHT OF NOMINAL VALUE 10 KG	29
S3	CALIBRATION OF A NOMINAL 10 KΩ STANDARD RESISTOR	31
S4	CALIBRATION OF A GAUGE BLOCK OF NOMINAL LENGTH 50 MM	34
S5	CALIBRATION OF A TYPE N THERMOCOUPLE AT 1000 °C	37
S6	CALIBRATION OF A POWER SENSOR AT A FREQUENCY OF 19 GH	42
S7	CALIBRATION OF A COAXIAL STEP ATTENUATOR AT A SETTING OF 30 DB (INCREMENTAL LOSS)	46
	SUPPLEMENT 2	51
S8	INTRODUCTION	52
S9	CALIBRATION OF A HAND-HELD DIGITAL MULTIMETER AT 100 V DC 55	
S10	CALIBRATION OF A VERNIER CALLIPER	59
S11	CALIBRATION OF A TEMPERATURE BLOCK CALIBRATOR AT A TEMPERATURE OF 180 °C	64
S12	CALIBRATION OF A HOUSEHOLD WATER METER	67
S13	CALIBRATION OF A RING GAUGE WITH A NOMINAL DIAMETER OF 90 MM⁷²	

1 INTRODUCTION

1.1 This document sets down the principles of and the requirements for the evaluation of the uncertainty of measurement in calibration and the statement of this uncertainty in calibration certificates based on the ILAC policy for uncertainty in calibration as stated in the ILAC P14 [ref. 5]. Both ILAC-P14 and EA-4/02 are mandatory for Accreditation Bodies that are EA members. The formulation is kept on a general level to suit all fields of calibration. The method outlined may have to be supplemented by more specific advice for different fields, to make the information more readily applicable. In developing such supplementary guidelines, the general principles stated in this document should be followed to ensure harmonisation between the different fields.

1.2 The formulation in this document is in accordance with JCGM 100:2008, Evaluation of measurement data – Guide to the Expression of uncertainty in measurement (GUM 1995 with minor correction). That document has been elaborated by the Joint Committee for Guides in Metrology, in which participate BIPM, IEC, IFCC, ILAC, ISO, IUPAC, IUPAP and OIML. [ref. 1]. But whereas [ref. 1] establishes general rules for evaluating and expressing uncertainty in measurement that can be followed in most fields of physical measurements, this document concentrates on the method most suitable for the measurements in calibration laboratories and describes an unambiguous and harmonised way of evaluating and stating the uncertainty of measurement. However, other approaches proposed by the GUM and its supplements (as for example the Monte Carlo method) are acceptable. It comprises the following subjects:

- definitions basic to the document;
- methods for evaluating the uncertainty of measurement of input quantities in a measurement model;
- relationship between the uncertainty of measurement of the output quantity and the evaluated uncertainty of the input quantities;
- expanded uncertainty of measurement of the output quantity in a measurement model;
- statement of the uncertainty of measurement;
- a step-by-step procedure for calculating the uncertainty of measurement.

Evaluation of uncertainty of calibration is also addressed in several of the EURAMET's calibration guidelines available at www.euramet.org

2 OUTLINE AND DEFINITIONS

Note: Terms of special relevance to the context of the main text are written in bold when they appear for the first time in this document. Appendix B contains a glossary of these terms together with references.

2.1 The statement of the result of a measurement is complete only if it contains both the value attributed to the measurand and the uncertainty of measurement associated with that value. In this document all quantities which are not exactly known are treated

as **random variables**, including the influence quantities which may affect the measured value.

2.2 The **uncertainty of measurement** is a non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used [ref. 3]. In this document the shorthand term **uncertainty** is used for **uncertainty of measurement** if there is no risk of misunderstanding. For typical sources of uncertainty in a measurement see the list given in Appendix C.

2.3 The **measurands** are the quantities intended to be measured. In calibration one usually deals with only one measurand or **output quantity** Y that depends upon a number of **input quantities** X_i ($i = 1, 2, \dots, N$) according to the functional relationship

$$Y = f(X_1, X_2, \dots, X_N) \quad (2.1)$$

The measurement function f represents a function of the input quantities, which, when calculated using known values for the input quantities, is a measured value of the output quantity.'

. It describes how values of the output quantity Y are obtained from values of the input quantities X_i . In most cases it will be an analytical expression, but it may also be a group of such expressions which include corrections and correction factors for systematic effects, thereby leading to a more complicated relationship that is not written down as one function explicitly. Further, f may be determined experimentally, or may exist only as a computer algorithm that must be evaluated numerically, or it may be a combination of any of these.

2.4 If they are independent, the set of input quantities X_i may be grouped into two categories according to the way in which the value of the quantity and its associated uncertainty have been determined:

- (a) quantities whose estimate and associated uncertainty are directly determined in the current measurement. These values may be obtained, for example, from a single observation, repeated observations, or judgement based on experience. They may involve the determination of corrections to instrument readings as well as corrections for influence quantities, such as ambient temperature, barometric pressure or humidity;
- (b) quantities whose estimate and associated uncertainty are brought into the measurement from external sources, such as quantities associated with calibrated measurement standards, certified reference materials or reference data obtained from handbooks.

If the input quantities are not independent, see clause 4.6.

2.5 An estimate of the measurand Y , the **output estimate** denoted by y , is obtained from equation (2.1) using **input estimates** x_i for the values of the input quantities X_i

$$y = f(x_1, x_2, \dots, x_N) \quad (2.2)$$

It is understood that the input values are best estimates that have been corrected for all effects significant for the model. If not, the necessary corrections have been introduced as separate input quantities.

- 2.6** For a random variable the **variance** of its distribution or the positive square root of the variance, called **standard deviation**, is used as a measure of the dispersion of its values. The **standard uncertainty of measurement** associated with the output estimate or measurement result y , denoted by $u(y)$, is the standard deviation of the measurand Y . It is to be determined from the estimates x_i of the input quantities X_i and their associated standard uncertainties $u(x_i)$.

The standard uncertainty associated with an estimate has the same unit as the estimate.

In some cases, the **relative standard uncertainty of measurement** may be appropriate which is the standard uncertainty of measurement associated with an estimate divided by the modulus of that estimate and is therefore dimensionless. This concept cannot be used if the estimate equals zero.

3 EVALUATION OF THE UNCERTAINTY OF MEASUREMENT OF INPUT ESTIMATES

General considerations

- 3.1.1 The uncertainty of measurement associated with the input estimates is evaluated according to either a 'Type A' or a 'Type B' method of evaluation. The **Type A evaluation of standard uncertainty** is the method of evaluating the uncertainty by the statistical analysis of a series of observations. In this case the standard uncertainty is the experimental standard deviation of the mean that follows from an averaging procedure or an appropriate regression analysis. The **Type B evaluation of standard uncertainty** is the method of evaluating the uncertainty by means other than the statistical analysis of a series of observations, typically by assigning an appropriate probability distribution function to the input quantity. In this case the evaluation of the standard uncertainty is based on some other scientific knowledge.

Note: There are occasions, seldom met in calibration, when all possible values of a quantity lie on one side of a single limit value. A well-known case is the so-called cosine error. For the treatment of such special cases, see [ref. 1].

Type A evaluation of standard uncertainty

- 3.2.1 Type A evaluation of standard uncertainty can be applied when several independent observations have been made for one of the input quantities under the same conditions of measurement. If there is sufficient resolution in the measurement process, there will be an observable scatter or spread in the values obtained.

- 3.2.2 Assume that the repeatedly measured input quantity X_i is the quantity Q . With n statistically independent observations ($n > 1$), the estimate of the quantity Q is \bar{q} , the **arithmetic mean** or the **average** of the individual observed values q_j ($j = 1, 2, \dots, n$)

$$\bar{q} = \frac{1}{n} \sum_{j=1}^n q_j \quad (3.1)$$

The uncertainty of measurement associated with the estimate \bar{q} is evaluated according to one of the following methods:

- (a) An estimate of the variance of the underlying probability distribution is the **experimental variance** $s^2(q)$ of values q_j that is given by

$$s^2(q) = \frac{1}{n-1} \sum_{j=1}^n (q_j - \bar{q})^2 \quad (3.2)$$

Its (positive) square root is termed **experimental standard deviation**. If the quantity reported is the arithmetic mean \bar{q} , the best estimate of the variance of the mean is the **experimental variance of the mean** given by

$$s^2(\bar{q}) = \frac{s^2(q)}{n} \quad (3.3)$$

Its (positive) square root is termed **experimental standard deviation of the mean**. The standard uncertainty $u(\bar{q})$ associated with the input estimate \bar{q} is the experimental standard deviation of the mean

$$u(\bar{q}) = s(\bar{q}) \quad (3.4)$$

Warning: Generally, when the number n of repeated measurements is low ($n < 10$), the reliability of a Type A evaluation of standard uncertainty, as expressed by equation (3.4), must be considered. If the number of observations cannot be increased, other means of evaluating the standard uncertainty given in the text must be considered.

- (b) For a measurement that is well-characterised and under statistical control a combined or **pooled estimate of variance** s_p^2 may be available that characterises the dispersion better than the estimated standard deviation obtained from a limited number of observations. If in such a case the value of the input quantity Q is determined as the arithmetic mean \bar{q} of a small number n of independent observations, the variance of the mean may be estimated by

$$s^2(\bar{q}) = \frac{s_p^2}{n} \quad (3.5)$$

The standard uncertainty is deduced from this value by equation (3.4).

Type B evaluation of standard uncertainty

- 3.3.1 Type B evaluation of standard uncertainty is the method of uncertainty evaluation associated with an estimate x_i of an input quantity X_i by means other than the statistical analysis of a series of observations. The standard uncertainty $u(x_i)$ is evaluated by scientific judgement based on all available information on the possible variability of X_i . Values belonging to this category may be derived from
- previous measurement data;
 - experience with or general knowledge of the behaviour and properties of relevant materials and instruments;
 - manufacturer's specifications;
 - data provided in calibration and other certificates;
 - uncertainties assigned to reference data taken from handbooks.
- 3.3.2 The proper use of the available information for a Type B evaluation of standard uncertainty of measurement calls for insight based on experience and general knowledge. It is a skill that can be learned with practice. A well-based Type B evaluation of standard uncertainty can be as reliable as a Type A evaluation of standard uncertainty, especially in a measurement situation where a Type A evaluation is based only on a comparatively small number of statistically independent observations. The following cases must be discerned:
- (a) When only a **single value** is known for the quantity X_i , e.g., a single measured value, a resultant value of a previous measurement, a reference value from the literature, or a correction value, this value will be used for x_i . The standard uncertainty $u(x_i)$ associated with x_i is to be adopted where it is given. Otherwise, it has to be calculated from unequivocal uncertainty data. If the number of observations cannot be increased, a different approach to estimation of the standard uncertainty given in b) has to be considered.
 - (b) When a **probability distribution** can be assumed for the quantity X_i , based on theory or experience, then the appropriate expectation or expected value and the square root of the variance of this distribution have to be taken as the estimate x_i and the associated standard uncertainty $u(x_i)$, respectively.
 - (c) If only **upper** and **lower limits** a_+ and a_- can be estimated for the value of the quantity X_i (e.g., manufacturer's specifications of a measuring instrument, a temperature range, a rounding or truncation error resulting from automated data reduction), a probability distribution with constant probability density between these limits (rectangular probability distribution) has to be assumed for the possible variability of the input quantity X_i . According to case (b) above this leads to

$$x_i = \frac{1}{2}(a_+ + a_-) \quad (3.6)$$

for the estimated value and

$$u^2(x_i) = \frac{1}{12}(a_+ - a_-)^2 \quad (3.7)$$

for the square of the standard uncertainty. If the difference between the limiting values is denoted by $2a$, equation (3.7) yields

$$u^2(x_i) = \frac{1}{3}a^2 \quad (3.8)$$

The rectangular distribution is a reasonable description in probability terms of one's inadequate knowledge about the input quantity X_i in the absence of any other information than its limits of variability. But if it is known that values of the quantity in question near the centre of the variability interval are more likely than values close to the limits, a triangular or normal distribution may be a better model. On the other hand, if values close to the limits are more likely than values near the centre, a U-shaped distribution may be more appropriate. For the evaluation of uncertainty in these cases see [ref. 1].

4 CALCULATION OF THE STANDARD UNCERTAINTY OF THE OUTPUT ESTIMATE

4.1 For uncorrelated input quantities the square of the standard uncertainty associated with the output estimate y is given by

$$u^2(y) = \sum_{i=1}^N u_i^2(y) \quad (4.1)$$

The quantity $u_i(y)$ ($i = 1, 2, \dots, N$) is the contribution to the standard uncertainty associated with the output estimate y resulting from the standard uncertainty associated with the input estimate x_i

$$u_i(y) = c_i u(x_i) \quad (4.2)$$

where c_i is the **sensitivity coefficient** associated with the input estimate x_i , i.e., the partial derivative of the model function f with respect to X_i , evaluated at the input estimates x_i ,

$$c_i = \frac{\partial f}{\partial x_i} = \left. \frac{\partial f}{\partial x_i} \right|_{X_1=x_1, X_2=x_2, \dots, X_n} \quad (4.3)$$

Note: There are cases, seldom occurring in calibration, where the model function is strongly non-linear or some of the sensitivity coefficients [see equation (4.2) and (4.3)] vanish and higher order terms have to be included into equation (4.1). For a treatment of such special cases see ref. 1 and Supplement Examples S4 and S13.

4.2 The sensitivity coefficient c_i describes the extent to which the output estimate y is influenced by variations of the input estimate x_i . It can be evaluated from the model function f by equation (4.3) experimentally, or by using numerical methods, i.e., by calculating the change in the output estimate y due to a corresponding change in the

input estimate x_i of $+u(x_i)$ and $-u(x_i)$ and taking as the value of c_i the resulting difference in y divided by $2u(x_i)$. Sometimes it may be more appropriate to find the change in the output estimate y from an experiment by repeating the measurement at e.g., $x_i \pm u(x_i)$.

4.3 Whereas $u(x_i)$ is always positive, the contribution $u_i(y)$ according to equation (4.2) is either positive or negative, depending on the sign of the sensitivity coefficient c_i . The sign of $u_i(y)$ has to be taken into account in the case of correlated input quantities, see equation (D4) of Appendix D.

4.4 If the model function f is a linear combination of the input quantities X_i with multiplicative factors p_i

$$f(X_1, X_2, \dots, X_N) = \sum_{i=1}^N p_i X_i \quad (4.4)$$

the output estimate according to equation (2.2) is given by the corresponding linear combination of the input estimates

$$y = \sum_{i=1}^N p_i x_i \quad (4.5)$$

whereas the sensitivity coefficients equal p_i and equation (4.1) converts to

$$u^2(y) = \sum_{i=1}^N p_i^2 u^2(x_i) \quad (4.6)$$

4.5 If the model function f is a product or quotient of the input quantities X_i raised to powers p_i

$$f(X_1, X_2, \dots, X_N) = c \prod_{i=1}^N X_i^{p_i} \quad (4.7)$$

the output estimate is the corresponding product or quotient of the input estimates

$$y = c \prod_{i=1}^N x_i^{p_i} \quad (4.8)$$

The sensitivity coefficients equal $p_i y/x_i$ in this case and an expression analogous to equation (4.6) is obtained from equation (4.1), if relative standard uncertainties $w(y) = u(y)/|y|$ and $w(x_i) = u(x_i)/|x_i|$ are used,

$$w^2(y) = \sum_{i=1}^N p_i^2 w^2(x_i) \quad (4.9)$$

4.6 If two input quantities X_i and X_k are **correlated** to some degree, i.e., if they are mutually dependent, their **covariance** also must be considered as a contribution to

the uncertainty. See Appendix D for how this must be done. The ability to take into account the effect of correlations depends on the knowledge of the measurement and on the judgement of mutual dependency of the input quantities. In general, neglecting correlations between input quantities can lead to an incorrect evaluation of the standard uncertainty of the measurand.

- 4.7** The covariance associated with the estimates of two input quantities X_i and X_k may be taken to be zero or treated as insignificant if
- the input quantities X_i and X_k are independent, for example, because they have been repeatedly but not simultaneously observed in different independent experiments or because they represent resultant quantities of different evaluations that have been made independently, or if
 - either of the input quantities X_i and X_k can be treated as constant, or if
 - investigation gives no information indicating the presence of correlation between the input quantities X_i and X_k .

Sometimes correlations can be eliminated by a proper choice of the model function. See, for example, Appendix D, para. D.6

- 4.8** The uncertainty analysis for a measurement — sometimes called the uncertainty budget of the measurement — should include a list of all sources of uncertainty together with the associated standard uncertainties of measurement and the methods of evaluating them. For repeated measurements the number n of observations also must be stated. For the sake of clarity, it is recommended to present the data relevant to this analysis in the form of a table. In this table all quantities should be referenced by a physical symbol X_i or a short identifier. For each of them at least the estimate x_i , the associated standard uncertainty of measurement $u(x_i)$, the sensitivity coefficient c_i and the different uncertainty contributions $u_i(y)$ should be specified. The measurement unit of each of the quantities should also be stated with the numerical values given in the table.

- 4.9** A formal example of such an arrangement is given as Table 4.1 applicable for the case of uncorrelated input quantities. The standard uncertainty associated with the measurement result $u(y)$ given in the bottom right corner of the table is the root sum square of all the uncertainty contributions in the outer right column. The grey part of the table is not completed.

Table 4.1: Schematic of an ordered arrangement of the quantities, estimates, standard uncertainties, sensitivity coefficients and uncertainty contributions used in the uncertainty analysis of a measurement.

Quantity	Estimate	Standard uncertainty	Probability distribution	Sensitivity coefficient	Contribution to the standard uncertainty
X_i	x_i	$u(x_i)$		c_i	$u_i(y)$
X_1	x_1	$u(x_1)$	<i>Prob. dist. 1</i>	c_1	$u_1(y)$
X_2	x_2	$u(x_2)$	<i>Prob. dist. 2</i>	c_2	$u_2(y)$
:	:	:		:	:
X_N	x_N	$u(x_N)$	<i>Prob. dist. N</i>	c_N	$u_N(y)$
Y	y				$u(y)$

5 EXPANDED UNCERTAINTY OF MEASUREMENT

- 5.1 Within EA it has been decided that calibration laboratories accredited by members of the EA shall state an **expanded uncertainty of measurement** U , obtained by multiplying the standard uncertainty $u(y)$ of the output estimate y by a **coverage factor** k ,

$$U = ku(y) \quad (5.1)$$

In cases where a normal (Gaussian) distribution can be attributed to the measurand and the standard uncertainty associated with the output estimate has sufficient reliability, the standard coverage factor $k = 2$ shall be used. The assigned expanded uncertainty corresponds to a **coverage probability** of approximately 95 %. These conditions are fulfilled in the majority of cases encountered in calibration work.

- 5.2 The assumption of a normal distribution cannot always be easily confirmed experimentally. However, in the cases where several (i.e., $N \geq 3$) uncertainty components, derived from well-behaved probability distributions of independent quantities, e.g., normal distributions or rectangular distributions, contribute to the standard uncertainty associated with the output estimate by comparable amounts, the conditions of the Central Limit Theorem are met, and it can be assumed to an adequate degree of approximation that the distribution of the output quantity is normal.
- 5.3 The reliability of the standard uncertainty assigned to the output estimate is determined by its effective degrees of freedom (see Appendix E). However, the reliability criterion is satisfied if none of the uncertainty contributions is obtained from a Type A evaluation based on fewer than ten repeated observations.
- 5.4 If one of these conditions (normality or sufficient reliability) is not fulfilled, the standard coverage factor $k = 2$ can yield an expanded uncertainty corresponding to a coverage probability of less than 95 %. In these cases, in order to ensure that a value of the expanded uncertainty is quoted corresponding to the same coverage probability as in the normal case, other procedures have to be followed. The use of approximately the same coverage probability is essential whenever two results of measurement of the same quantity have to be compared, e.g., when evaluating the results of an inter-laboratory comparison or assessing compliance with a specification.
- 5.5 Even if a normal distribution can be assumed, it may still occur that the standard uncertainty associated with the output estimate is of insufficient reliability. If, in this case, it is not expedient to increase the number n of repeated measurements or to use a Type B evaluation instead of the Type A evaluation of poor reliability, the method given in Appendix E should be used.
- 5.6 For the remaining cases, i.e., all cases where the assumption of a normal distribution cannot be justified, information on the actual probability distribution of the output estimate must be used to obtain a value of the coverage factor k that corresponds to a coverage probability of approximately 95 %.
- The Monte Carlo method of GUM Supplement 1 can be applied in such cases.

6 STEP-BY-STEP PROCEDURE FOR CALCULATING THE UNCERTAINTY OF MEASUREMENT

6.1 The following is a guide to the use of this document in practice (cf. worked examples in the supplement):

- (a) Express in mathematical terms the dependence of the measurand (output quantity) Y on the input quantities X_i according to equation (2.1). In the case of a direct comparison of two standards the equation may be very simple, e.g., $Y = X_1 - X_2$.
- (b) Identify and apply all significant corrections.
- (c) List all sources of uncertainty in the form of an uncertainty analysis in accordance with Section 4.
- (d) Calculate the standard uncertainty $u(\bar{q})$ for repeatedly measured quantities in accordance with sub-section 3.2.
- (e) For single values, e.g., resultant values of previous measurements, correction values or values from the literature, adopt the standard uncertainty where it is given or can be calculated according to paragraph 3.3.2 (a). If no data are available from which the standard uncertainty can be derived, state a value of $u(x_i)$ on the basis of scientific experience.
- (f) For input quantities for which the probability distribution is known or can be assumed, calculate the expectation and the standard uncertainty $u(x_i)$ according to paragraph 3.3.2 (b). If only upper and lower limits are given or can be estimated, calculate the standard uncertainty $u(x_i)$ in accordance with paragraph 3.3.2 (c).
- (g) Calculate for each input quantity X_i the contribution $u_i(y)$ to the standard uncertainty associated with the output estimate resulting from the input estimate x_i according to equations (4.2) and (4.3) and sum their squares as described in equation (4.1) to obtain the square of the standard uncertainty $u(y)$ of the measurand. If input quantities are known to be correlated, apply the procedure given in Appendix D.
- (h) Calculate the expanded uncertainty U by multiplying the standard uncertainty $u(y)$ associated with the output estimate by a coverage factor k chosen in accordance with Section 5.
- (i) Report the result of the measurement comprising the estimate y of the measurand, the associated expanded uncertainty U and the coverage factor k in the calibration certificate in accordance with Section 6 of ILAC P14 [ref. 5].

Note: In some circumstances the acceptability of calibration results is to be decided by comparing them against a specification. The decision is made according to an agreed Decision Rule which states how to take account of the associated uncertainty of measurement. Appendix F gives some brief guidance on the subject.

7 REFERENCES

- [1] JCGM 100:2008 GUM 1995 with minor corrections, Evaluation of measurement data – Guide to the expression of uncertainty in measurement. (Available from www.bipm.org)
- [2] ISO/IEC 17025:2017, General requirements for the competence of testing and calibration laboratories.
- [3] JCGM 200:2012 International vocabulary of metrology – Basic and general concepts and associated terms (Available from www.bimp.org)
- [4] International Standard ISO 3534-1, Statistics-Vocabulary and symbols-Part 1: General statistical terms and terms used in probability – (ISO 3534-1:2006)
- [5] ILAC-P14:09/2020 – ILAC Policy for Measurement Uncertainty in Calibration
- [6] JCGM 104:2009 Evaluation of measurement data – An introduction to the “Guide to the expression of uncertainty in measurement” and related documents. (Available from www.bipm.org)
- [7] JCGM 106 (ISO/IEC Guide 98-4:2012) Uncertainty of measurement -- Part 4: Role of measurement uncertainty in conformity assessment
- [8] ILAC G8:2019 Guidelines on Decision Rules and Statements of Conformity
- [9] ISO 10576-1:2003 Statistical methods – Guidelines for the evaluation of conformity with specified requirements
- [10] ASME B89.7.4.1-2005 Measurement Uncertainty and Conformance Testing: Risk Analysis

APPENDIX A
CALIBRATION AND MEASUREMENT CAPABILITY

The concept of a calibration and measurement capability, CMC, is thoroughly investigated in the paper on calibration and measurement capabilities issued by the joint BIPM/ILAC Working Group 7 September 2007. This paper is included in the ILAC policy for uncertainty in calibration as an annex and the policy is the basis for a harmonised approach to the CMC among accredited laboratories around the world [ref.5].

The methods for evaluation of uncertainty outlined in this document should be used when accredited laboratories establish their CMC.

APPENDIX B

GLOSSARY OF SOME RELEVANT TERMS

- B1 arithmetic mean** ([ref.1] Section C.2.19)
Average; The sum of values divided by the number of values.
- B2 calibration and measurement capability**
The calibration and measurement capability (CMC) is expressed in terms of:
1. Measurand or reference material;
 2. Calibration/measurement method/procedure and/or type of instrument/material to be calibrated/measured;
 3. Measurement range and additional parameters where applicable, e.g., frequency of applied voltage;
 4. Uncertainty of measurement.
- For a complete explanation see ref.5.
- B3 correlation coefficient** (from [ref. 1] Section C.3.6)
The correlation coefficient is a measure of the relative mutual dependence of two variables, equal to the ratio of their covariances to the positive square root of the product of their variances. For a more elaborated description see ref.1.
- B4 covariance** (from [ref. 1] Section C.3.4)
The measure of the mutual dependence of two random variables, equal to the expectation of the product of the deviations of two random variables from their respective expectations. The complete definition can be found in ref.1.
- B5 coverage factor** ([ref. 3] term 2.38)
Number larger than one by which a combined standard measurement uncertainty is multiplied to obtain an expanded measurement uncertainty.
- B6 coverage probability** ([ref. 3] term 2.37)
Probability that the set of true quantity values of a measurand is contained within a specified coverage interval.
Note: The term “true value” (see Appendix D) is not used in this Guide for the reasons given in D.3.5 of the GUM; the terms “value of a measurand” (or of a quantity) and “true value of a measurand” (or of a quantity) are viewed as equivalent. (GUM 3.1.1)
See also ref.6 (JCGM 104:2009) chapter 1.
- B7 experimental standard deviation** ([ref. 1] section 4.2.2)
The positive square root of the experimental variance.
- B8 expanded (measurement) uncertainty** ([ref. 3] term 2.35)
Product of a combined standard measurement uncertainty and a factor larger than the number one.
- B9 experimental variance** (from [ref. 1] Section 4.2.2)
The quantity that characterises the dispersion of the results of a series of n observations of the same measurand given by equation (3.2) in the text.

- B10 input estimate** (from [ref. 1] Section 4.1.4 and C2.26)
The estimate of an input quantity used in the evaluation of the result of a measurement.
- B11 input quantity** (from [ref. 1] Section 4.1.2)
A quantity on which the measurand depends, taken into account in the process of evaluating the result of a measurement.
- B12 measurand** ([ref. 3] term 2.3)
Quantity intended to be measured.
- B13 measurement uncertainty**, uncertainty of measurement, uncertainty ([ref.3] Section 2.26)
Non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used.
- B14 output estimate** (from [ref. 1] Section 4.1.4 and C2.26)
The result of a measurement calculated from the input estimates by the model function.
- B15 output quantity** (from [ref. 1] Section 4.1.4)
The quantity that represents the measurand in the evaluation of a measurement result.
- B16 pooled estimate of variance** (from [ref. 1] Section 4.2.4)
An estimate of the experimental variance obtained from long series of observations of the same measurand in well-characterised measurements under statistical control.
- B17 probability distribution** ([ref. 1] Section C.2.3)
A function giving the probability that a random variable takes any given value or belongs to a given set of values.
- B18 random variable** ([ref. 1] Section C.2.2)
A variable that may take any of the values of a specified set of values and with which is associated a probability distribution.
- B19 relative standard uncertainty of measurement** ([ref. 3] Section 2.32)
Standard measurement uncertainty divided by the absolute value of the measured quantity value.
- B20 sensitivity coefficient associated with an input estimate** (from [ref. 1] Section 5.1.3)
The differential change in the output estimate generated by a differential change in an input estimate divided by the change in that input estimate.
- B21 standard deviation** ([ref. 1] Section C.2.12)
The positive square root of the variance.

- B22 standard measurement uncertainty** ([ref. 3] term 2.30)
The measurement uncertainty expressed as a standard deviation.
- B23 Type A evaluation of measurement uncertainty** ([ref. 3] Section 2.28)
Estimation of a component of measurement uncertainty by a statistical analysis of measured quantity values obtained under defined measurement conditions.
- B24 Type B evaluation of measurement uncertainty** ([ref. 3] term 2.29)
Estimation of a component of measurement uncertainty determined by means other than a Type A evaluation of measurement uncertainty.
- B25 uncertainty budget** ([ref. 3] Section 2.33)
Statement of a measurement uncertainty, of the components of that measurement uncertainty, and of their calculation and combination.
- B26 variance** (from [ref. 1] Section C.2.11)
The expectation of the square of the centred random variable.

APPENDIX C

SOURCES OF UNCERTAINTY OF MEASUREMENT

- C1** The uncertainty of the result of a measurement reflects the lack of complete knowledge of the value of the measurand. Complete knowledge requires an infinite amount of information. Phenomena that contribute to the uncertainty and thus to the fact that the result of a measurement cannot be characterised by a unique value, are called sources of uncertainty. In practice, there are many possible sources of uncertainty in a measurement [ref. 1], including:
- (a) incomplete definition of the measurand;
 - (b) imperfect realisation of the definition of the measurand;
 - (c) non-representative sampling — the sample measured may not represent the defined measurand;
 - (d) inadequately known effects of environmental conditions or imperfect measurements of these;
 - (e) personal bias in reading analogue instruments;
 - (f) finite instrument resolution or discrimination threshold;
 - (g) inexact values of measurement standards and reference materials;
 - (h) inexact values of constants and other parameters obtained from external sources and used in the data-reduction algorithm;
 - (i) approximations and assumptions incorporated in the measurement method and procedure;
 - (j) variations in repeated observations of the measurand under apparently identical conditions.
- C2** These sources are not necessarily independent. Some of the sources (a) to (i) may contribute to (j).
- C3** Most examples in Supplements 1 and 2 show the handling of uncertainty contributions (g), typically the value and uncertainty from reference standards, (b), the changes in the value of the reference since calibration e.g., due to drift or instrument specifications, (d) unknowns with respect to environmental influence factors and (j) variability in the readings or partial results obtained.

APPENDIX D

CORRELATED INPUT QUANTITIES

- D1** If two input quantities X_i and X_k are known to be correlated— i.e., if they are dependent on each other in one way or another — the **covariance** associated with the two estimates x_i and x_k

$$u(x_i, x_k) = u(x_i)u(x_k)r(x_i, x_k) \quad (i \neq k) \quad (D.1)$$

has to be considered as an additional contribution to the uncertainty. The degree of correlation is characterised by the **correlation coefficient** $r(x_i, x_k)$ (where $i \neq k$ and $|r| \leq 1$).

- D2** In the case of n independent pairs of simultaneously repeated observations of two quantities P and Q the covariance associated with the arithmetic means \bar{p} and \bar{q} is given by

$$s(\bar{p}, \bar{q}) = \frac{1}{n(n-1)} \sum_{j=1}^n (p_j - \bar{p})(q_j - \bar{q}) \quad (D.2)$$

and by substitution of s from equation (D.2) for $u(x_i, x_k)$ in equation (D.1), r can be calculated from equation (D.1).

- D3** For influence quantities correlation can be based on experience. When there is correlation, equation (4.1) must be replaced by

$$u^2(y) = \sum_{i=1}^N c_i^2 u^2(x_i) + 2 \sum_{i=1}^{N-1} \sum_{k=i+1}^N c_i c_k u(x_i, x_k) \quad (D.3)$$

where c_i and c_k are the sensitivity coefficients defined by equation (4.3) or

$$u^2(y) = \sum_{i=1}^N u_i^2(y) + 2 \sum_{i=1}^{N-1} \sum_{k=i+1}^N u_i(y) u_k(y) r(x_i, x_k) \quad (D.4)$$

with the contributions $u_i(y)$ to the standard uncertainty of the output estimate y resulting from the standard uncertainty of the input estimate x_i given by equation (4.2). It should be noted that the second summation of terms in equation (D.3) or (D.4) may be negative in sign.

- D4** In practice, input quantities are often correlated because the same physical reference standard, measuring instrument, reference datum, or even measurement method having a significant uncertainty is used in the evaluation of these quantities. Without loss of generality, suppose that two input quantities X_1 and X_2 estimated by x_1 and x_2 depend on the set of independent variables Q_l ($l = 1, 2, \dots, L$)

$$\begin{aligned} X_1 &= g_1(Q_1, Q_2, \dots, Q_L) \\ X_2 &= g_2(Q_1, Q_2, \dots, Q_L) \end{aligned} \quad (D.5)$$

although some of these variables may not necessarily appear in both functions. The estimates x_1 and x_2 of the input quantities will be correlated to some extent, even if the estimates q_l ($l = 1, 2, \dots, L$) are uncorrelated. In that case the covariance $u(x_1, x_2)$ associated with the estimates x_1 and x_2 is given by

$$u(x_1, x_2) = \sum_{l=1}^L c_{1l} c_{2l} u^2(q_l) \quad (D.6)$$

where c_{1i} and c_{2i} are the sensitivity coefficients derived from the functions g_1 and g_2 in analogy to equation (4.3). Because only those terms contribute to the sum for which the sensitivity coefficients do not vanish, the covariance is zero if no variable is common to functions g_1 and g_2 . The correlation coefficient $r(x_1, x_2)$ associated with the estimates x_1 and x_2 is determined from equation (D.6) together with equation (D.1).

- D5** The following example demonstrates correlations which exist between values attributed to two artefact standards that are calibrated against the same reference standard.

Measurement Problem

Two standards characterized by quantities X_1 and X_2 are compared with a reference standard characterized by the quantity Q_S by means of a measuring system capable of determining a difference z in their values with an associated standard uncertainty $u(z)$. The value q_S of the reference standard is known with standard uncertainty $u(q_S)$.

Mathematical Model

The estimates x_1 and x_2 depend on the value q_S of the reference standard and the observed differences z_1 and z_2 according to the relations

$$\begin{aligned} x_1 &= q_S - z_1 \\ x_2 &= q_S - z_2 \end{aligned} \tag{D.7}$$

Standard uncertainties and covariances

Suppose the quantities X_1 , X_2 and Q_S can be assumed to be independent. The standard uncertainties are calculated from equation (4.4) and the covariance associated with the estimates x_1 and x_2 is calculated from equation (D.6), assuming that $u(z_1) = u(z_2) = u(z)$,

$$\begin{aligned} u^2(x_1) &= u^2(q_S) + u^2(z) \\ u^2(x_2) &= u^2(q_S) + u^2(z) \\ u(x_1, x_2) &= u^2(q_S) \end{aligned} \tag{D.8}$$

The correlation coefficient deduced from these results is

$$r(x_1, x_2) = \frac{u^2(q_S)}{u^2(q_S) + u^2(z)} \tag{D.9}$$

Its value ranges from 0 to +1 depending on the ratio of the standard uncertainties $u(q_S)$ and $u(z)$.

- D6** The case described by equation (D.5) is an occasion where the inclusion of correlation in the evaluation of the standard uncertainty of the measurand can be avoided by a careful choice of variable in choice of the model function. Introducing directly the independent variables Q_i by replacing the original variables X_1 and X_2 in the model function f in accordance with the transformation equations (D.5) gives a new model function that does not contain the correlated variables X_1 and X_2 any longer.
- D7** There are cases however, where correlation between two input quantities X_1 and X_2 cannot be avoided, e.g., using the same measuring instrument or the same reference

standard when determining the input estimates x_1 and x_2 but where transformation equations to new independent variables are not available.

If furthermore the degree of correlation is not exactly known it may be useful to assess the maximum influence this correlation can have by an upper bound estimate of the standard uncertainty of the measurand which in the case that other correlations have not to be taken into account takes the form

$$u^2(y) \leq \left(|u_1(y)| + |u_2(y)| \right)^2 + u_r^2(y) \quad (\text{D.10})$$

with $u_r(y)$ being the contribution to the standard uncertainty of all the remaining input quantities assumed to be uncorrelated.

Note: Equation (D.10) is easily generalised to cases of one or several groups with two or more correlated input quantities. In this case a respective worst-case sum must be introduced into equation (D.10) for each group of correlated quantities.

APPENDIX E

COVERAGE FACTORS DERIVED FROM EFFECTIVE DEGREES OF FREEDOM

- E1** To determine the value of a coverage factor k corresponding to a specified coverage probability requires that the reliability of the standard uncertainty $u(y)$ of the output estimate y is taken into account. That means taking into account how well $u(y)$ estimates the standard deviation associated with the result of the measurement. For an estimate of the standard deviation of a normal distribution, the degrees of freedom of this estimate, which depends on the size of the sample on which it is based, is a measure of the reliability. Similarly, a suitable measure of the reliability of the standard uncertainty associated with an output estimate is its effective degrees of freedom ν_{eff} , which is approximated by an appropriate combination of the effective degrees of freedom of its different uncertainty contributions $u_i(y)$.
- E2** The procedure for calculating an appropriate coverage factor k when the conditions of the Central Limit Theorem are met comprises the following three steps:
- Obtain the standard uncertainty associated with the output estimate according to the step-by-step procedure given in Section 7.
 - Determine the effective degrees of freedom ν_{eff} of the standard uncertainty $u(y)$ associated with the output estimate y from the Welch-Satterthwaite formula.

$$\nu_{\text{eff}} = \frac{u^4(y)}{\sum_{i=1}^N \frac{u_i^4(y)}{\nu_i}}, \quad (\text{E.1})$$

where the $u_i(y)$ ($i = 1, 2, \dots, N$), defined in equation (4.2), are the contributions to the standard uncertainty associated with the output estimate y resulting from the standard uncertainty associated with the input estimate x_i which are assumed to be mutually statistically independent, and ν_i is the effective degrees of freedom of the standard uncertainty contribution $u_i(y)$.

For a standard uncertainty $u(q)$ obtained from a Type A evaluation as discussed in sub-section 3.1, the degrees of freedom are given by $\nu_i = n-1$. It is more problematic to associate degrees of freedom with a standard uncertainty $u(x_i)$ obtained from a Type B evaluation. However, it is common practice to carry out such evaluations in a manner that ensures that any underestimation is avoided. If, for example, lower and upper limits a_- and a_+ are set, they are usually chosen in such a way that the probability of the quantity in question lying outside these limits is in fact extremely small. Under the assumption that this practice is followed, the degrees of freedom of the standard uncertainty $u(x_i)$ obtained from a Type B evaluation may be taken to be *infinite*.

- (c) Obtain the coverage factor k from the table of values given as Table E.1 of this Appendix. This table is based on a t -distribution evaluated for a coverage probability of 95,45%. If ν_{eff} is not an integer, which will usually be the case, truncate ν_{eff} to the next lower integer.

Table E.1: Coverage factors k for different effective degrees of freedom ν_{eff} for approximately 95 % coverage (95.45 %)

ν_{eff}	1	2	3	4	5	6	7	8	9	10
k	13,97	4,53	3,31	2,87	2,65	2,52	2,43	2,37	2,32	2,28
ν_{eff}	11	12	13	14	15	16	17	18	19	20
k	2,25	2,23	2,21	2,20	2,18	2,17	2,16	2,15	2,14	2,13
ν_{eff}	25	30	35	40	45	50	∞			
k	2,11	2,09	2,07	2,06	2,06	2,05	2,00			

APPENDIX F
CONFORMITY WITH A SPECIFICATION

F1 In some circumstances the acceptability of calibration results is decided by comparing them against a specification, usually expressed in terms of a tolerance interval or limits. In these circumstances a Decision Rule is employed which dictates how the measurement uncertainty shall be taken into account when reaching such a conformity decision [ref. 7].

F2 For limit-based specifications, the area of the Probability Density Function (PDF) for a measurand Y that falls within the tolerance interval C represents the proportion of conforming values of the measurand that could be responsible for the measured value y_m . This is the conformance probability p_c

$$p_c = \int_C g(\eta; y_m, u_m) d\eta$$

F3 For example, assuming normal distribution, in the figure F1 the unshaded region of the PDF is within the tolerance interval and represents conforming values of the measurand that can be associated with the measurement result.




Figure F1: A measurement result within a tolerance interval that is defined by a single upper limit

T_u = upper limit; T_L = lower limit; u_m = uncertainty; U_m = expanded uncertainty; y_m = measured value

F4 The shaded region in the figure represents non-conforming values of the measurand that can similarly also be attributed to the measurement result. If it was decided to Accept the result, based upon the observation that the measured value y_m is inside the tolerance interval, the area of the shaded region corresponds to the Probability of False Acceptance (PFA). Similar arguments are used to define the Probability of False Rejection (PFR) when a decision is made to reject a result, based upon an observation that the measured value y_m falls outside a tolerance interval.

- F5** Decision Rules are commonly expressed in terms of conformance probability, in terms of the risk associated with an incorrect decision (PFA and PFR) or in other terms that relate the uncertainty to the size of the tolerance interval. They can have multiple outcomes, such as {Pass, Conditional Pass, Conditional Fail, Fail} or {Pass, Uncertain, Fail} or a simple binary {Pass, Fail} outcome, depending on how the Decision Rule dictates that uncertainty should be taken into account.
- F6** Examples of such rules and associated conformity decisions can be found in reference documents.

SUPPLEMENT 1

Examples

S1 INTRODUCTION

- S1.1** The following examples are chosen to demonstrate the method of evaluating the uncertainty of measurement. More typical and representative examples based on appropriate models have to be developed by special working groups in the different areas. Nevertheless, the examples presented here provide general guidance on how to proceed.
- S1.2** The examples are based on drafts prepared by EA Expert Groups. These drafts have been simplified and harmonised to make them transparent to laboratory staff in all fields of calibration. It is thus hoped that this set of examples will contribute to a better understanding of the details of setting up the model of evaluation (the measurement model) and to the harmonisation of the process of evaluating the uncertainty of measurement, independent of the field of calibration.
- S1.3** The contributions and values given in the examples are not intended to imply mandatory or preferred requirements. Laboratories should determine the uncertainty contributions on the basis of the model function they use in the evaluation of the particular calibration they perform and report the evaluated uncertainty of measurement on the calibration certificate they issue. In the examples given, the conditions stated in section 5 for the use of the standard coverage factor $k = 2$ are generally fulfilled.
- S1.4** The presentation of the examples follows, in accordance with the step-by-step procedure of section 6 of EA-4/02, a common scheme containing:
- a short descriptive title,
 - a general description of the process of measurement,
 - the model of evaluation with a list of symbols used,
 - an extended listing of input data with short descriptions of how they have been obtained,
 - the list of observations and the evaluation of the statistical parameters,
 - an uncertainty budget in table form,
 - the expanded uncertainty of measurement,
 - the reported complete result of measurement.
- S1.5** This first supplement to EA-4/02 is intended to be followed by others containing further worked-out examples on the evaluation of uncertainty of measurement in connection with the calibration of instruments. Examples may also be found in EURAMET Calibration Guides dealing with the calibration of specific types of measurement instruments.

S2 CALIBRATION OF A WEIGHT OF NOMINAL VALUE 10 KG

S2.1 The calibration of a weight of nominal value 10 kg of OIML class M1 is carried out by comparison to a reference standard (OIML class F2) of the same nominal value using a mass comparator whose performance characteristics have previously been determined.

S2.2 The unknown conventional mass m_X is obtained from:

$$m_X = m_S + \delta m_D + \delta m + \delta m_C + \delta B \quad (\text{S2.1})$$

where:

m_S - conventional mass of the standard,

δm_D - drift of value of the standard since its last calibration,

δm - observed difference in mass between the unknown mass and the standard,

δm_C - correction for eccentricity and magnetic effects,

δB - correction for air buoyancy.

S2.3 Reference standard (m_S): The calibration certificate for the reference standard gives a value of 10 000,005 g with an associated expanded uncertainty of 45 mg (coverage factor $k = 2$).

S2.4 Drift of the value of the standard (δm_D): The drift of the value of the reference standard is estimated from previous calibrations to be zero within ± 15 mg.

S2.5 Comparator ($\delta m, \delta m_C$): A previous evaluation of the repeatability of the mass difference between two weights of the same nominal value gives a pooled estimate of standard deviation of 25 mg. No correction is applied for the comparator, whereas variations due to eccentricity and magnetic effects are estimated to have rectangular limits of ± 10 mg.

S2.6 Air buoyancy (δB): No correction is made for the effects of air buoyancy, the limits of deviation are estimated to be $\pm 1 \times 10^{-6}$ of the nominal value assuming that the weights are of similar material.

S2.7 Correlation: None of the input quantities is considered to be correlated to any significant extent.

S2.8 Measurements: Three sets of observations of the difference in mass between the unknown mass and the standard are obtained using the substitution method and the substitution scheme ABBA ABBA ABBA:

Note: The substitution scheme ABBA means that, for each set of observations, the standard mass is first measured, then the unknown mass is measured twice, and finally the standard mass is measured.

no	conventional mass		Reading (g)		observed difference (g)
1	standard		+0,010		
	unknown		+0,020		
	unknown		+0,025		
	standard		+0,015		+0,010
2	standard		+0,025		
	unknown		+0,050		
	unknown		+0,055		
	standard		+0,020		+0,030
3	standard		+0,025		
	unknown		+0,045		
	unknown		+0,040		
	standard		+0,020		+0,020

arithmetic mean: $\overline{\delta m} = 20 \text{ mg}$

pooled estimate of standard deviation:
(obtained from prior evaluation) $s_p(\delta m) = 25 \text{ mg}$

standard uncertainty: $u(\delta m) = s(\overline{\delta m}) = \frac{25 \text{ mg}}{\sqrt{3}} = 14,4 \text{ mg}$

S2.9 Uncertainty budget (m_X):

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
m_S	10 000,005 g	22,5 mg	normal	1,0	22,5 mg
δm_D	0,000 g	8,66 mg	rectangular	1,0	8,66 mg
δm	0,020 g	14,4 mg	normal	1,0	14,4 mg
δm_C	0,000 g	5,77 mg	rectangular	1,0	5,77 mg
δB	0,000 g	5,77 mg	rectangular	1,0	5,77 mg
m_X	10 000,025 g				29,2 mg

S2.10 Expanded uncertainty

$$U = k \times u(m_X) = 2 \times 29,2 \text{ mg} \cong 58 \text{ mg}$$

Although the determination of δm is based on only three repeated readings, the evaluated uncertainty relies on a previously determined pooled estimate of the standard deviation. Thus the effective degrees of freedom is large, and together with the fact that the dominant uncertainty contributor has a normal distribution, the standard coverage factor, $k=2$, can be used.

S2.11 Reported result

The measured mass of the nominal 10 kg weight is 10,000 025 kg \pm 58 mg. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S3 CALIBRATION OF A NOMINAL 10 K Ω STANDARD RESISTOR

S3.1 The resistance of a four-terminal standard resistor is determined by direct substitution using a long-scale digital multimeter (7½ digit DMM) on its resistance range, and a calibrated four-terminal standard resistor of the same nominal value as the item to be calibrated as reference standard. The resistors are immersed in a well stirred oil bath operating at a temperature of 23 °C monitored by a centrally placed mercury-in-glass thermometer. The resistors are allowed to stabilise before the measurement. The four-terminal connectors of each resistor are connected in turn to the terminals of the DMM. It is determined that the measuring current of 100 μ A on the 10 k Ω range of the DMM is sufficiently low not to cause any appreciable self-heating of the resistors. The measurement procedure used also ensures that the effects of external leakage resistances on the result of measurement can be considered to be insignificant.

S3.2 The resistance R_X of the unknown resistor is obtained from the relationship:

$$R_X = (R_S + \delta R_D + \delta R_{TS}) r_C r - \delta R_{TX} \quad (\text{S3.1})$$

where:

- R_S - resistance of the reference,
- δR_D - drift of the resistance of the reference since its last calibration,
- δR_{TS} - temperature related resistance variation of the reference,
- $r = R_{IX}/R_S$ - ratio of the indicated resistance (index i means 'indicated') for the unknown and reference resistors,
- r_C - correction factor for parasitic voltages and instrument resolution
- δR_{TX} - temperature-related resistance variation of the unknown resistor.

S3.3 Reference standard (R_S): The calibration certificate for the reference standard gives a resistance value of 10 000,053 $\Omega \pm$ 5 m Ω (coverage factor $k = 2$) at the specified reference temperature of 23 °C.

- S3.4 Drift of the value of the standard (δR_D):** The drift of the resistance of the reference resistor since its last calibration is estimated from its calibration history to be +20 m Ω with deviations within ± 10 m Ω .
- S3.5 Temperature corrections (δR_{TS} , δR_{TX}):** The temperature of the oil bath is monitored using a calibrated thermometer to be 23,00 °C. Taking into account the metrological characteristics of the thermometer used and of gradients of temperature within the oil bath, the temperature of the resistors is estimated to coincide with the monitored temperature within $\pm 0,055$ K. Thus the known value 5×10^{-6} K $^{-1}$ of the temperature coefficient (TC) of the reference resistor gives limits $\pm 2,75$ m Ω for the deviation from its resistance value according to calibration, due to a possible deviation from the operating temperature. From the manufacturer's literature, the TC of the unknown resistor is estimated not to exceed $1,0 \times 10^{-5}$ K $^{-1}$, thus the resistance variation of the unknown resistor due to a temperature variation is estimated to be within $\pm 5,5$ m Ω .
- S3.6 Resistance measurements (r_C):** Since the same DMM is used to observe both R_{IX} and R_{IS} the uncertainty contributions are correlated but the effect is to reduce the uncertainty and it is only necessary to consider the relative difference in the resistance readings due to systematic effects such as parasitic voltages and instrument resolution (see the mathematical note in paragraph S3.12), which are estimated to have limits of $\pm 0,5 \times 10^{-6}$ for each reading. The distribution resulting for the ratio r_C is triangular with expectation 1,000 000 0 and limits $\pm 1,0 \times 10^{-6}$.
- S3.7 Correlation:** None of the input quantities is considered to be correlated to any significant extent.
- S3.8 Measurements (r):** Five observations are made to record the ratio r .

No.	observed ratio
1	1,000 010 4
2	1,000 010 7
3	1,000 010 6
4	1,000 010 3
5	1,000 010 5

arithmetic mean: $\bar{r} = 1,000\ 010\ 5$

experimental standard deviation: $s(r) = 0,158 \times 10^{-6}$

standard uncertainty: $u(r) = s(\bar{r}) = \frac{0,158 \times 10^{-6}}{\sqrt{5}} = 0,0707 \times 10^{-6}$

S3.9 Uncertainty budget (R_x):

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
R_s	10 000,053 Ω	2,5 m Ω	normal	1,0	2,5 m Ω
δR_D	0,020 Ω	5,8 m Ω	rectangular	1,0	5,8 m Ω
δR_{TS}	0,000 Ω	1,6 m Ω	rectangular	1,0	1,6 m Ω
δR_{TX}	0,000 Ω	3,2 m Ω	rectangular	1,0	3,2 m Ω
r_C	1,000 000 0	$0,41 \times 10^{-6}$	triangular	10 000 Ω	4,1 m Ω
r	1,000 010 5	$0,07 \times 10^{-6}$	normal	10 000 Ω	0,7 m Ω
R_x	10 000,178 Ω				8,33 m Ω

S3.10 Expanded uncertainty:

$$U = k \times u(R_x) = 2 \times 8,33 \text{ m}\Omega \cong 17 \text{ m}\Omega$$

S3.11 Reported result: The measured value of the nominal 10 k Ω resistor, at a measuring temperature of 23,00 $^{\circ}\text{C}$ and a measuring current of 100 μA , is (10 000,178 \pm 0,017) Ω .

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S3.12 Mathematical note on the standard uncertainty of measurement of the ratio of indicated resistance values: The unknown and the reference resistors have nearly the same resistance. Within the usual linear approximation in the deviations, the values causing the DMM indications R_{iX} and R_{iS} are given by

$$\begin{aligned} R_X' &= R_{iX} \left(1 + \frac{\delta R_X'}{R}\right) \\ R_S' &= R_{iS} \left(1 + \frac{\delta R_S'}{R}\right) \end{aligned} \quad (\text{S3.2})$$

with R being the nominal value of the resistors and $\delta R_X'$ and $\delta R_S'$ the unknown deviations. The resistance ratio deduced from these expressions is

$$\frac{R_X'}{R_S'} = r r_C \quad (\text{S3.3})$$

with the ratio of the indicated resistance for the unknown and the reference resistor

$$r = \frac{R_{iX}}{R_{iS}} \quad (\text{S3.4})$$

and the correction factor (linear approximation in the deviations)

$$r_C = 1 + \frac{\delta R_X' - \delta R_S'}{R} \quad (\text{S3.5})$$

Because of the fact that the difference of the deviations enters into equation (S3.5), correlated contributions of systematic effects resulting from the internal scale of the DMM do not influence the result. The standard uncertainty of the correction factor is determined only by uncorrelated deviations resulting from the parasitic effects and the resolution of the DMM. Assuming that $u(\delta R_x') = u(\delta R_s') = u(\delta R')$, it is given by the expression

$$u^2(r_c) = 2 \frac{u^2(\delta R')}{R^2} \quad (\text{S3.6})$$

S4 CALIBRATION OF A GAUGE BLOCK OF NOMINAL LENGTH 50 MM

S4.1 The calibration of the grade 0 gauge block (ISO 3650) of 50 mm nominal length is carried out by comparison using a comparator and a calibrated gauge block of the same nominal length and the same material as reference standard. The difference in central length is determined in vertical position of the two gauge blocks using two length indicators contacting the upper and lower measuring faces. The actual length l_x' of the gauge block to be calibrated is related to the actual length l_s' of the reference standard by the equation

$$l_x' = l_s' + \delta l \quad (\text{S4.1})$$

with δl being the measured length difference. l_x' and l_s' are the lengths of the gauge blocks under measurement conditions, in particular at a temperature which, on account of the uncertainty in the measurement of laboratory temperature, may not be identical with the reference temperature for length measurements.

S4.2 The length l_x of the unknown gauge block at the reference temperature is obtained from the relationship:

$$l_x = l_s + \delta l_D + \delta l + \delta l_C - L(\bar{\alpha} \times \delta t + \delta \alpha \times \Delta \bar{t}) - \delta l_V \quad (\text{S4.2})$$

where:

- l_s - length of the reference gauge block at the reference temperature $t_0 = 20^\circ\text{C}$ according to its calibration certificate;
- δl_D - change of the length of the reference gauge block since its last calibration due to drift;
- δl - observed difference in length between the unknown and the reference gauge block;
- δl_C - correction for non-linearity and offset of the comparator;
- L - nominal length of the gauge blocks considered;
- $\bar{\alpha} = (\alpha_x + \alpha_s) / 2$ - average of the thermal expansion coefficients of the unknown and reference gauge blocks;
- $\delta t = (t_x - t_s)$ - temperature difference between the unknown and reference gauge blocks;

- $\delta\alpha = (\alpha_x - \alpha_s)$ - difference in the thermal expansion coefficients between the unknown and the reference gauge blocks;
- $\Delta\bar{t} = (t_x + t_s)/2 - t_0$ - deviation of the average temperature of the unknown and the reference gauge blocks from the reference temperature;
- δl_v - correction for non-central contacting of the measuring faces of the unknown gauge block.

- S4.3 Reference standard (l_s):** The length of the reference gauge block together with the associated expanded uncertainty of measurement is given in the calibration certificate of a set of gauge blocks as 50,000 02 mm \pm 30 nm (coverage factor $k = 2$).
- S4.4 Drift of the standard (δl_b):** The temporal drift of the length of the reference gauge block is estimated from previous calibrations to be zero with limits ± 30 nm. General experience with gauge blocks of this type suggests that zero drift is most probable and that a triangular probability distribution can be assumed.
- S4.5 Comparator (δl_c):** The comparator has been verified to meet the specifications stated in EURAMET cg-2 Version 2.0 (03/2011) (previously EAL-G21:1996). From this, it can be ascertained that for length differences D up to ± 10 μ m corrections to the indicated length difference are within the limits $\pm(30 \text{ nm} + 0,02x|D|)$, where D has unit μ m. Taking into account the tolerances of the grade 0 gauge block to be calibrated and the grade K reference gauge block, the maximum length difference will be within ± 1 μ m leading to limits of ± 32 nm for non-linearity and offset corrections of the comparator used.
- S4.6 Temperature corrections ($\bar{\alpha}$, δt , $\delta\alpha$, $\Delta\bar{t}$):** Before calibration, care is taken to ensure that the gauge blocks assume ambient temperature of the measuring room. The remaining difference in temperature between the standard and the gauge block to be calibrated is estimated to be within $\pm 0,05$ K. Based on the calibration certificate of the reference gauge block and the manufacturer's data for the gauge block to be calibrated the linear thermal expansion coefficient of the steel gauge blocks is assumed to be within the interval $(11,5 \pm 1,0) \times 10^{-6} \text{ }^\circ\text{C}^{-1}$. Combining the two rectangular distributions the difference in linear thermal expansion coefficient is triangularly distributed within the limits $\pm 2 \times 10^{-6} \text{ }^\circ\text{C}^{-1}$. The deviation of the mean temperature of measurement from the reference temperature $t_0 = 20$ $^\circ\text{C}$ is estimated to be within $\pm 0,5$ $^\circ\text{C}$. The best estimates of the difference in linear expansion coefficients and the deviations of the mean temperature from the reference temperature are zero. Therefore second order terms have to be taken into account in the evaluation of their uncertainty contribution resulting in the product of standard uncertainties associated with the factors of the product term $\delta\alpha \times \Delta\bar{t}$ in equation (S4.2) (see the mathematical note in paragraph S4.13, eq. (S4.5)). The final standard uncertainty is $u(\delta\alpha \times \Delta\bar{t}) = 0,236 \times 10^{-6}$.

S4.7 Variation in length (δl): For gauge blocks of grade 0 the variation in length determined from measurements at the centre and the four corners has to be within $\pm 0,10 \mu\text{m}$ (ISO 3650:1998). Assuming that this variation occurs on the measuring faces along the short edge of length 9 mm and that the central length is measured inside a circle of radius 0,5 mm, the correction due to central misalignment of the contacting point is estimated to be within $\pm 6,7 \text{ nm}$.

S4.8 Correlation: None of the input quantities is considered to be correlated to any significant extent.

S4.9 Measurements (δl): The following observations are made for the difference between the unknown gauge block and the reference standard, the comparator being reset using the reference standard before each reading.

obs. no.	obs. value
1	-100 nm
2	-95 nm
3	-80 nm
4	-95 nm
5	-100 nm

arithmetic mean:

$$\overline{\delta l} = -94 \text{ nm}$$

pooled estimate of standard deviation:

$$s_p(\delta l) = 12 \text{ nm}$$

(obtained from prior evaluation)

standard uncertainty:

$$u(\delta l) = s(\overline{\delta l}) = \frac{12 \text{ nm}}{\sqrt{5}} = 5,37 \text{ nm}$$

The pooled estimate of the standard deviation has been taken from the tests made to confirm compliance of the comparator used with the requirements of EURAMET cg-2 Version 2.0 (03/2011) (previously EAL-G21:1996).

S4.10 Uncertainty budget (δk):

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
l_s	50,000 020 mm	15 nm	normal	1,0	15,0 nm
δl_b	0 mm	12,2 nm	triangular	1,0	12,2 nm
δl	-0,000 094 mm	5,37 nm	normal	1,0	5,37 nm
δl_c	0 mm	18,5 nm	rectangular	1,0	18,5 nm
δt	0 °C	0,0289 °C	rectangular	-575 nm°C ⁻¹	-16,6 nm
$\delta \alpha \times \Delta \bar{t}$	0	$0,236 \times 10^{-6}$	special	-50 mm	-11,8 nm
δl_v	0 mm	3,87 nm	rectangular	-1,0	-3,87 nm
l_k	49,999 926 mm				34.3 nm

S4.11 Expanded uncertainty

$$U = k \times u(l_x) = 2 \times 34.3 \text{ nm} \cong 69 \text{ nm}$$

S4.12 Reported result

The measured value of the nominal 50 mm gauge block is 49,999 926 mm \pm 69 nm. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S4.13 Mathematical note on the standard uncertainty of measurement of the product of two quantities with zero expectation:

If a product of two quantities is considered, the usual method of evaluation of uncertainty contributions based on the linearisation of the model function has to be modified if one or both of the expectations of the factors in the product are zero. If the factors in the product are statistically independent with non-zero expectations, the square of the relative standard uncertainty of measurement (relative variance) associated with the product can be expressed without any linearisation by the squares of the relative standard uncertainties associated with the estimates of the factors:

$$w^2(x_1 \times x_2) = w^2(x_1) + w^2(x_2) + w^2(x_1) \times w^2(x_2) \quad (\text{S4.2})$$

Using the definition of the relative standard uncertainty of measurement this expression is easily transformed into the general relation

$$u^2(x_1 \times x_2) = x_2^2 u^2(x_1) + x_1^2 u^2(x_2) + u^2(x_1) \times u^2(x_2) \quad (\text{S4.3})$$

If the standard uncertainties $u(x_1)$ and $u(x_2)$ associated with the expectations x_1 and x_2 are much smaller than the moduli of the respective expectation values the third term on the right side may be neglected. The resulting equation represents the case described by the usual method based on the linearisation of the model function.

If, however, one of the moduli of the expectation values, for example $|x_2|$, is much smaller than the standard uncertainty $u(x_2)$ associated with this expectation or even zero, the product term involving this expectation may be neglected on the right side of equation (S4.3), but not the third term. The resulting equation is

$$u^2(x_1 \times x_2) \cong x_1^2 u^2(x_2) + u^2(x_1) \times u^2(x_2) \quad (\text{S4.4})$$

If both moduli of the expectation values are much smaller than their associated standard uncertainties or even zero, only the third term in equation (S4.3) gives a significant contribution:

$$u^2(x_1 \times x_2) \cong u^2(x_1) \times u^2(x_2) \quad (\text{S4.5})$$

Another example of a contribution which not always can be subject to reasonable linearisation, is $Y = X^2$. If X is normally distributed with mean m and variance s^2 , then the expectation value of Y is $m^2 + s^2$ and the standard uncertainty is $\sqrt{2(2m^2 s^2 + s^4)}$. For example, if the X has mean $m = 0$, $Y = X^2$ should be assigned the value $y = s^2$ with $u(y) = \sqrt{2} s^2$

S5 CALIBRATION OF A TYPE N THERMOCOUPLE AT 1000 °C

S5.1 A type N thermocouple is calibrated by comparison with two reference thermocouples of type R in a horizontal furnace at a temperature of 1000 °C. The emfs generated by

the thermocouples are measured using a digital voltmeter through a selector/reversing switch. All thermocouples have their reference junctions at 0 °C. The thermocouple to be calibrated is connected to the reference point using compensating cables. Temperature values are given in the t_{90} temperature scale.

S5.2 The temperature t_x of the hot junction of the thermocouple to be calibrated is

$$t_x = t_s(V_{IS} + \delta V_{IS1} + \delta V_{IS2} + \delta V_R - \frac{\delta t_{0S}}{C_{S0}}) + \delta t_D + \delta t_F$$

$$\cong t_s(V_{IS}) + C_S \times \delta V_{IS1} + C_S \times \delta V_{IS2} + C_S \times \delta V_R - \frac{C_S}{C_{S0}} \delta t_{0S} + \delta t_D + \delta t_F$$
(S5.1)

S5.3 The voltage V_x across the thermocouple wires with the cold junction at 0 °C during calibration is

$$V_x(t) \cong V_x(t_x) + \frac{\Delta t}{C_x} - \frac{\delta t_{0X}}{C_{X0}}$$

$$= V_{IX} + \delta V_{IX1} + \delta V_{IX2} + \delta V_R + \delta V_{LX} + \frac{\Delta t}{C_x} - \frac{\delta t_{0X}}{C_{X0}}$$
(S5.2)

where:

- $t_s(V)$ - temperature of the reference thermometer in terms of voltage with cold junction at 0 °C. The function is given in the calibration certificate;
- V_{IS}, V_{IX} - indication of the voltmeter;
- $\delta V_{IS1}, \delta V_{IX1}$ - voltage corrections obtained from the calibration of the voltmeter;
- $\delta V_{IS2}, \delta V_{IX2}$ - voltage corrections due to the limited resolution of the voltmeter;
- δV_R - voltage correction due to contact effects of the reversing switch;
- $\delta t_{0S}, \delta t_{0X}$ - temperature corrections due to the deviation of the reference temperatures from 0 °C;
- C_S, C_x - sensitivities of the thermocouples for voltage at the measuring temperature of 1000 °C;
- C_{S0}, C_{X0} - sensitivities of the thermocouples for voltage at the reference temperature of 0 °C;
- δt_D - change of the values of the reference thermometers since their last calibration due to drift;
- δt_F - temperature correction due to non-uniformity of the temperature of the furnace;
- t - temperature at which the thermocouple is to be calibrated (calibration point);

- $\Delta t = t - t_x$ - deviation of the temperature of the calibration point from the temperature of the furnace;
- δV_{Lx} - voltage correction due to the compensating cables.

S5.4 The reported result is the output emf of the thermocouple at the temperature of its hot junction. Because the measurement process consists of two steps — determination of the temperature of the furnace and determination of emf of the thermocouple to be calibrated — the evaluation of the uncertainty of measurement is split in two parts.

S5.5 Reference standards ($t_s(V)$): The reference thermocouples are supplied with calibration certificates that relate the temperature at their hot junction with their cold junction at 0 °C to the voltage across their wires. The expanded uncertainty of measurement at 1000 °C is $U = 0,3$ °C (coverage factor $k = 2$).

S5.6 Calibration of the voltmeter ($\delta V_{IS1}, \delta V_{IX1}$): The voltmeter has been calibrated. Corrections to the measured voltages are made to all results. The calibration certificate gives a constant expanded uncertainty of measurement for voltages smaller than 50 mV of $U = 2,0$ μ V (coverage factor $k = 2$).

S5.7 Resolution of the voltmeter ($\delta V_{IS2}, \delta V_{IX2}$): A 4½ digit microvoltmeter has been used in its 10 mV range resulting in resolution limits of $\pm 0,5$ μ V at each indication.

S5.8 Parasitic voltages (δV_R): Residual parasitic offset voltages due to the switch contacts have been estimated to be zero within ± 2 μ V.

S5.9 Reference temperatures ($\delta t_{0S}, \delta t_{0X}$): The temperature of the reference point of each thermocouple is known to be 0 °C within $\pm 0,1$ °C.

S5.10 Voltage sensitivities (C_S, C_X, C_{S0}, C_{X0}): The voltage sensitivities of the thermocouples have been taken from reference tables:

	1000 °C	0 °C
reference thermocouple	$C_S = 0,077$ °C/ μ V	$C_{S0} = 0,189$ °C/ μ V
unknown thermocouple	$C_X = 0,026$ °C/ μ V	$C_{X0} = 0,039$ °C/ μ V

S5.11 Drift of the reference standard (δt_D): From previous calibrations the drift of the reference standards are estimated to be zero within the limits $\pm 0,3$ °C.

S5.12 Temperature gradients (δt_F): The temperature gradients inside the furnace have been measured. At 1000 °C, deviations from non-uniformity of temperature in the region of measurement are within ± 1 °C.

S5.13 Compensating cables (δV_{Lx}): The compensating cables have been investigated in the range 0 °C to 40 °C. From this, the voltage differences between the cables and the thermocouple wires are estimated to be within ± 5 μ V.

S5.14 Measurements (V_{is} , $t_s(V_{is})$, V_{ix}): The indications of the voltmeter are recorded in the following operational procedure which gives four readings for every thermocouple and reduces the effects of temperature drift in the thermal source and of parasitic thermal voltages in the measuring circuit:

1st cycle:

1st standard, unknown thermocouple, 2nd standard,
2nd standard, unknown thermocouple, 1st standard.

Reversion of polarity.

2nd cycle:

1st standard, unknown thermocouple, 2nd standard,
2nd standard, unknown thermocouple, 1st standard.

S5.15 The procedure requires that the difference between the two reference standards must not exceed $\pm 0,3$ °C. If the difference is not within these limits the observations have to be repeated and/or the reasons for such a large difference have to be investigated.

Thermocouple	1 st reference	Unknown	2 nd reference
Indicated voltage, corrected	+10 500 μV	+36 245 μV	+10 503 μV
	+10 503 μV	+36 248 μV	+10 503 μV
	-10 503 μV	-36 248 μV	-10 505 μV
	-10 504 μV	-36 251 μV	-10 505 μV
Mean voltage	10 502,5 μV	36 248 μV	10 504 μV
Temperature of the hot junction	10 00,4 °C		1000,6 °C
Temperature of the furnace		1000,5 °C	

S5.16 From the four readings on each thermocouple given in the table above, the mean value of the voltages of each thermocouple is deduced. The voltage values of the reference thermocouples are converted into temperature values by means of the temperature-voltage relations stated in their calibration certificates. The observed temperature values are highly correlated (correlation factor nearly one). Therefore, by taking their mean value, they are combined to one observation only, which is the temperature of the furnace at the location of the thermocouple to be calibrated. In a similar way, one observation of the voltage of the thermocouple to be calibrated has been extracted. In order to evaluate the uncertainty of measurement associated with these observations, a series of ten measurements has been previously undertaken at the same temperature of operation. It gave a pooled estimate of standard deviation for the temperature of the furnace and the voltage of the thermocouple to be calibrated.

The respective standard uncertainties of measurement of the observed quantities are:

pooled estimate of standard deviation: $s_p(t_S) = 0,10 \text{ }^\circ\text{C}$
 standard uncertainty: $u(t_S) = \frac{s_p(t_S)}{\sqrt{1}} = 0,10 \text{ }^\circ\text{C}$
 pooled estimate of standard deviation: $s_p(V_{IX}) = 1,6 \text{ } \mu\text{V}$
 standard uncertainty: $u(V_{IX}) = \frac{s_p(V_{IX})}{\sqrt{1}} = 1,6 \text{ } \mu\text{V}$

S5.17 Uncertainty budget (temperature t_x of the furnace):

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
t_S	1000,5 $^\circ\text{C}$	0,10 $^\circ\text{C}$	normal	1,0	0,10 $^\circ\text{C}$
δV_{IS1}	0 μV	1,00 μV	normal	0,077 $^\circ\text{C}/\mu\text{V}$	0,077 $^\circ\text{C}$
δV_{IS2}	0 μV	0,29 μV	rectangular	0,077 $^\circ\text{C}/\mu\text{V}$	0,022 $^\circ\text{C}$
δV_R	0 μV	1,15 μV	rectangular	0,077 $^\circ\text{C}/\mu\text{V}$	0,089 $^\circ\text{C}$
δt_{0S}	0 $^\circ\text{C}$	0,058 $^\circ\text{C}$	rectangular	-0,407	-0,023 $^\circ\text{C}$
δt_S	0 $^\circ\text{C}$	0,15 $^\circ\text{C}$	normal	1,0	0,15 $^\circ\text{C}$
δt_D	0 $^\circ\text{C}$	0,173 $^\circ\text{C}$	rectangular	1,0	0,173 $^\circ\text{C}$
δt_F	0 $^\circ\text{C}$	0,577 $^\circ\text{C}$	rectangular	1,0	0,577 $^\circ\text{C}$
t_x	1000,5 $^\circ\text{C}$				0,641 $^\circ\text{C}$

S5.18 Uncertainty budget (emf V_x of the thermocouple to be calibrated):

The standard uncertainty of measurement associated with the temperature deviation of the calibration point from the temperature of the furnace is the standard uncertainty of measurement associated with the temperature of the furnace because the temperature point is a defined value (exactly known).

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
V_{IX}	36 248 μV	1,60 μV	normal	1,0	1,60 μV
δV_{IX1}	0 μV	1,00 μV	normal	1,0	1,00 μV
δV_{IX2}	0 μV	0,29 μV	rectangular	1,0	0,29 μV
δV_R	0 μV	1,15 μV	rectangular	1,0	1,15 μV
δV_{LX}	0 μV	2,9 μV	rectangular	1,0	2,9 μV
Δt	0,5 $^\circ\text{C}$	0,641 $^\circ\text{C}$	normal	38,5 $\mu\text{V}/^\circ\text{C}$	24,5 μV
δt_{0X}	0 $^\circ\text{C}$	0,058 $^\circ\text{C}$	rectangular	-25,6 $\mu\text{V}/^\circ\text{C}$	-1,48 μV
V_x	36 229 μV				25,0 μV

S5.19 Expanded uncertainties

The expanded uncertainty associated with the measurement of the temperature of the furnace is

$$U = k \times u(t_x) = 2 \times 0,641 \text{ }^\circ\text{C} \cong 1,3 \text{ }^\circ\text{C}$$

The expanded uncertainty associated with the emf value of the thermocouple to be calibrated is

$$U = k \times u(V_x) = 2 \times 25,0 \text{ } \mu\text{V} \cong 50 \text{ } \mu\text{V}$$

S5.20 Reported result


The type N thermocouple shows, at the temperature of 1000,0 °C with its cold junction at a temperature of 0 °C, an emf of 36 230 μV ± 50 μV.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor *k* such that the coverage probability corresponds to approximately 95 %.

S6 CALIBRATION OF A POWER SENSOR AT A FREQUENCY OF 19 GHZ

S6.1 The measurement involves the calibration of an unknown power sensor with respect to a calibrated power sensor used as a reference by substitution on a stable transfer standard of known small reflection coefficient. The measurement is made in terms of calibration factor, which is defined as the ratio of incident power at the reference frequency of 50 MHz to the incident power at the calibration frequency under the condition that both incident powers give equal power sensor response. At each frequency, one determines the (indicated) ratio of the power for the sensor to be calibrated, respectively the reference sensor and the internal sensor that forms part of the transfer standard, using a dual power meter with ratio facility.

S6.2 Schematic of the measuring system



S6.3 The quantity K , termed ‘calibration factor’ by some manufacturers, is defined as:

$$K = \frac{P_{\text{Ir}}}{P_{\text{Ic}}} = \frac{(1 + |\Gamma_{\text{r}}|^2) P_{\text{Ar}}}{(1 + |\Gamma_{\text{c}}|^2) P_{\text{Ac}}} \quad (\text{S6.1})$$

for the equal power meter indication

where:

- P_{Ir} - incident power at the reference frequency (50 MHz),
- P_{Ic} - incident power at the calibration frequency,
- Γ_{r} - voltage reflection coefficient of the sensor at the reference frequency
- Γ_{c} - voltage reflection coefficient of the sensor at the calibration frequency
- P_{Ar} - power absorbed by the sensor at the reference frequency
- P_{Ac} - power absorbed by the sensor at the calibration frequency

S6.4 The calibration factor of the unknown sensor is obtained from the relationship

$$K_{\text{X}} = (K_{\text{S}} + \delta K_{\text{D}}) \frac{M_{\text{Sr}} M_{\text{Xc}}}{M_{\text{Sc}} M_{\text{Xr}}} p_{\text{Cr}} p_{\text{Cc}} p \quad (\text{S6.2})$$

where:

- K_{S} - calibration factor of the reference power sensor;
- δK_{D} - change of the calibration factor of the reference power sensor since its last calibration due to drift;
- M_{Sr} - mismatch factor of reference sensor at the reference frequency;
- M_{Sc} - mismatch factor of standard sensor at the calibration frequency;
- M_{Xr} - mismatch factor of sensor to be calibrated at the reference frequency;
- M_{Xc} - mismatch factor of sensor to be calibrated at the calibration frequency;
- p_{Cr} - correction of the observed ratio for non-linearity and limited resolution of the power meter at power ratio level of the reference frequency;
- p_{Cc} - correction of the observed ratio for non-linearity and limited resolution of the power meter at power ratio level of the calibration frequency;
- $p = \frac{p_{\text{Sr}} p_{\text{Xc}}}{p_{\text{Sc}} p_{\text{Xr}}}$ - observed ratio of power ratios derived from:
 - p_{Sr} - indicated power ratio for the reference sensor at the reference frequency;
 - p_{Sc} - indicated power ratio for the reference sensor at the calibration frequency;

- ρ_{Xr} - indicated power ratio for the sensor to be calibrated at the reference frequency;
- ρ_{Xc} - indicated power ratio for the sensor to be calibrated at the calibration frequency.

S6.5 Reference sensor (K_S): The reference sensor was calibrated six months before the calibration of the unknown power sensor. The value of the calibration factor, given in the calibration certificate, is $(95,7 \pm 1,1) \%$ (coverage factor $k = 2$), which may also be expressed as $0,957 \pm 0,011$.

S6.6 Drift of the standard (δK_D): The drift of the calibration factor of the reference standard is estimated from annual calibrations to be $-0,002$ per year with deviations within $\pm 0,004$. From these values, the drift of the reference sensor, which was calibrated half a year ago, is estimated to equal $-0,001$ with deviations within $\pm 0,002$.

S6.7 Linearity and resolution of the power meter (ρ_{Cr}, ρ_{Cc}): The expanded uncertainty of $0,002$ (coverage factor $k = 2$) is assigned to the power meter readings at the power ratio level of the reference frequency and of $0,0002$ (coverage factor $k = 2$) at the power ratio level of calibration frequency due to non-linearity of the power meter used. These values have been obtained from previous measurements. Since the same power meter has been used to observe both ρ_S and ρ_X , the uncertainty contributions at the reference as well at the calibration frequency are correlated. Because power ratios at both frequencies are considered, the effect of the correlations is to reduce the uncertainty. Thus, only the relative difference in the readings due to systematic effects should be taken into account (see the mathematical note in paragraph S3.12), resulting in a standard uncertainty of $0,00142$ associated with the correction factor ρ_{Cr} and $0,000142$ with the correction factor ρ_{Cc} .

The expanded uncertainty of measurement stated for the readings of the power meter contains linearity and resolution effects. The linearity effects are correlated whereas the resolution effects are uncorrelated. As shown in S3.12, building the power ratio cancels the influence of correlations and gives a reduced standard uncertainty of measurement to be associated with the ratio. In the calculations above, however, the separated correlated and uncorrelated contributions are not known and the values given are upper bounds for the standard uncertainty of measurement associated with ratios. The uncertainty budget finally shows that the contributions arising from these ratios are insignificant, i.e. the approximations are justified.

S6.8 Mismatch factors ($M_{Sr}, M_{Sc}, M_{Xr}, M_{Xc}$): As the transfer standard system is not perfectly matched and the phase of the reflection coefficients of the transfer standard, the unknown and the standard power sensors are not known, there will be an uncertainty due to mismatch for each sensor at the reference frequency and at the calibration frequency. The corresponding limits of deviation have to be calculated for the reference and the calibration frequencies from the relationship:

$$M_{S,X} = 1 \pm 2|\Gamma_G||\Gamma_{S,X}| \quad (\text{S6.3})$$

where the magnitudes of the reflection coefficients of the transfer standard, the reference sensor and the sensor to be calibrated are:

	50 MHz	18 GHz
$ \Gamma_G $	0,02	0,07
$ \Gamma_S $	0,02	0,10
$ \Gamma_X $	0,02	0,12

The probability distribution of the individual contributions is U-shaped. This is taken into account by replacing the factor 1/3 for a rectangular distribution by 1/2 in calculating the variance from the square of the half-width determined from the limits. The standard uncertainty due to mismatch is therefore obtained from:

$$u(M_{s,x}) = \frac{2 |\Gamma_G| |\Gamma_{S,x}|}{\sqrt{2}} \quad (\text{S6.4})$$

Note: The values of the reflection coefficients are the results of measurements which are themselves subject to uncertainty. This is accounted for by adding the square root of the sum of the uncertainty of measurement squared and the measured value squared.

S6.9 Correlation: None of the input quantities is considered to be correlated to any significant extent.

S6.10 Measurements (p): Three separate readings are made which involve disconnection and reconnection of both the reference sensor and the sensor to be calibrated on the transfer standard to take connector repeatability into account. The power meter readings used to calculate the observed power ratio p are as follows:

obs. no	p_{Sr}	p_{Sc}	p_{Xr}	p_{Xc}	p
1	1,000 1	0,992 4	1,000 1	0,969 8	0,977 2
2	1,000 0	0,994 2	1,000 0	0,961 5	0,967 1
3	0,999 9	0,995 3	1,000 1	0,979 2	0,983 6

arithmetic mean: $\bar{p} = 0,976 0$

experimental standard deviation: $s(p) = 0,0083$

standard uncertainty: $u(p) = s(\bar{p}) = \frac{0,0083}{\sqrt{3}} = 0,0048$

S6.11 Uncertainty budget (K_x):

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
K_S	0,957	0,005 5	normal	0,976	0,0053 7
δK_D	-0,001	0,001 2	rectangular	0,976	0,0011 3
M_{Sr}	1,000	0,000 6	U-shaped	0,933	0,0005 3
M_{Sc}	1,000	0,009 9	U-shaped	-0,933	0,0092 4
M_{Xr}	1,000	0,000 6	U-shaped	-0,933	-0,0005 3
M_{Xc}	1,000	0,011 9	U-shaped	0,933	0,0111 0
p_{Cr}	1,000	0,001 4	normal	0,933	0,0013 2
p_{Cc}	1,000	0,000 1	normal	0,933	0,0001 3
p	0,976	0,004 8	normal	0,956	0,0045 9
K_x	0,933				0,0162 3

S6.12 Expanded uncertainty:

$$U = k \times u(K_x) = 2 \times 0,016\ 23 \cong 0,032$$

Although the determination of p is based on only three values (and thus $\nu = 2$), the contribution for p to the total uncertainty is about 1/4, thus the effect of the poor quality of the estimate of u_p is reduced. Hence, the standard coverage factor $k=2$ can be used."

In fact, there should be a calculation of ν_{eff} in the budget according to Appendix E ($\nu_{\text{eff}} \sim 310$)

S6.13 Reported result:

The calibration factor of the power sensor at 18 GHz is $0,933 \pm 0,032$, which may also be expressed as $(93,3 \pm 3,2) \%$.


The reported input quantities are considered expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S7 CALIBRATION OF A COAXIAL STEP ATTENUATOR AT A SETTING OF 30 DB (INCREMENTAL LOSS)

S7.1 The measurement involves the calibration of a coaxial step attenuator at 10 GHz using an attenuation measuring system containing a calibrated step attenuator which acts as the attenuation reference. The method of measurement involves the

determination of the attenuation between matched source and matched load. In this case the unknown attenuator can be switched between settings of 0 dB and 30 dB and it is this change (called incremental loss) that is determined in the calibration process. The attenuation measuring system has a digital readout and an analogue null detector which is used to indicate the balance condition.

S7.2 Schematic of the measuring system



S7.3 The attenuation L_x of the attenuator to be calibrated is obtained from the relation:

$$L_x = L_S + \delta L_S + \delta L_D + \delta L_M + \delta L_K + \delta L_{ib} - \delta L_{ia} + \delta L_{0b} - \delta L_{0a} \quad (S7.1)$$

where:

- $L_S = L_{ib} - L_{ia}$ - attenuation difference of reference attenuator derived from:
- L_{ia} - indicated attenuation with the attenuator to be calibrated, set at 0 dB;
- L_{ib} - indicated attenuation with the attenuator to be calibrated, set at 30 dB;
- δL_S - correction obtained from the calibration of the reference attenuator;
- δL_D - change of the attenuation of the reference attenuator since its last calibration due to drift;
- δL_M - correction due to mismatch loss;
- δL_K - correction for leakage signals between input and output of the attenuator to be calibrated due to imperfect isolation;
- $\delta L_{ia}, \delta L_{ib}$ - corrections due to the limited resolution of the reference detector at 0 dB and 30 dB settings;
- $\delta L_{0a}, \delta L_{0b}$ - corrections due to the limited resolution of the null detector at 0 dB and 30 dB settings.

S7.4 Reference attenuator (δL_S): The calibration certificate for the reference attenuator gives a value of attenuation for the 30,000 dB setting at 10 GHz of 30,003 dB with an associated expanded uncertainty of 0,005 dB (coverage factor $k = 2$). The correction of +0,003 dB with the associated expanded uncertainty of 0,005 dB (coverage factor $k = 2$) is considered to be valid for attenuation settings of the reference attenuator that differ not more than $\pm 0,1$ dB from the calibrated setting of 30,000 dB.

S7.5 Drift of the reference (δL_D): The drift of the attenuation of the reference attenuator is estimated from its calibration history to be zero with limits $\pm 0,002$ dB.

S7.6 Mismatch loss (δL_M): The reflection coefficients of the source and the load at the insertion point of the attenuator to be calibrated have been optimised by impedance matching to as low magnitudes as possible. Their magnitudes and the magnitudes of the scattering coefficients of the attenuator to be calibrated have been measured but their phase remains unknown. Without any phase information, a correction for mismatch error cannot be made, but the standard uncertainty (in dB) due to the incomplete knowledge of the match is estimated from the relationship [1]:

$$u(\delta L_M) = \frac{8,686}{\sqrt{2}} \sqrt{|T_S|^2 (|S_{11a}|^2 + |S_{11b}|^2) + |T_L|^2 (|S_{22a}|^2 + |S_{22b}|^2) + |T_S|^2 \times |T_L|^2 (|S_{21a}|^4 + |S_{21b}|^4)}$$
(S7.2)

with the source and load reflection coefficients

$$T_L = 0,03 \text{ and } T_S = 0,03$$

and the scattering coefficients of the attenuator to be calibrated at 10 GHz

	0 dB	30 dB
S_{11}	0,05	0,09
S_{22}	0,01	0,01
S_{21}	0,95	0,031

as $u(\delta L_M) = 0,020$ dB.

Note: The values of scattering and reflection coefficients are the results of measurements which are themselves not exactly known. This is accounted for by adding the square root of the sum of uncertainty of measurement squared and the measured value squared.

S7.7 Leakage correction (δL_K): Leakage signals through the attenuator to be calibrated have been estimated from the measurements at 0 dB setting to be at least 100 dB below the measurement signal. The correction for leakage signals is estimated from these findings to be within $\pm 0,003$ dB at the 30 dB setting.

S7.8 Resolution of the reference attenuator setting (δL_{ia} , δL_{ib}): The digital readout of the reference attenuator has a resolution of 0,001 dB from which the correction for resolution is estimated to be within $\pm 0,0005$ dB.

S7.9 Resolution of the null detector (δL_{0a} , δL_{0b}): The detector resolution was determined from a previous evaluation to have a standard deviation of 0,002 dB at each reading with assumed normal probability distribution.

S7.10 Correlation: None of the input quantities is considered to be correlated to any significant extent.

S7.11 Measurements (L_S): Four observations are made of the incremental loss of the attenuator to be calibrated between settings of 0 dB and 30 dB:

obs. no.	obs. values at	
	0 dB setting	30 dB setting
1	0,000 dB	30,033 dB
2	0,000 dB	30,058 dB
3	0,000 dB	30,018 dB
4	0,000 dB	30,052 dB

arithmetic mean: $\bar{L}_S = 30,040$ dB

experimental standard deviation: $s(L_S) = 0,018$ dB

standard uncertainty: $u(L_S) = s(\bar{L}_S) = \frac{0,018 \text{ dB}}{\sqrt{4}} = 0,009$ dB

S7.12 Uncertainty budget (L_X):

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
L_S	30,040 dB	0,0090 dB	normal	1,0	0,0090 dB
δL_S	0,003 dB	0,0025 dB	rectangular	1,0	0,0025 dB
δL_D	0 dB	0,0012 dB	rectangular	1,0	0,0012 dB
δL_M	0 dB	0,0198 dB	U-shaped	1,0	0,0198 dB
δL_K	0 dB	0,0017 dB	rectangular	1,0	0,0017 dB
δL_{ia}	0 dB	0,0003 dB	rectangular	-1,0	-0,0003 dB
δL_{ib}	0 dB	0,0003 dB	rectangular	1,0	0,0019 dB
δL_{0a}	0 dB	0,0020 dB	normal	-1,0	0,0020 dB
δL_{0b}	0 dB	0,0020 dB	normal	1,0	-0,0020 dB
L_X	30,043 dB				0,0223 dB

S7.13 Expanded uncertainty:

$$U = k \times u(L_X) = 2 \times 0,0223 \text{ dB} \cong 0,0446 \text{ dB}$$

The combined standard uncertainty u_c is about 2.5 times u_{L_S} , the only contribution with a limited degrees of freedom, $\nu = n - 1 = 3$. Hence $\nu_{\text{eff}} \sim 3 \times (2.5)^4 \sim 100$, and the use of $k=2$ is warranted.

S7.14 Reported result:

The measured value of the step attenuator for a setting of 30 dB at 10 GHz is $(30,043 \pm 0,045)$ dB.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S7.15 Reference

- [1] Harris, I. A. ; Warner, F. L. : *Re-examination of mismatch uncertainty when measuring microwave power and attenuation*. In: IEE Proc., Vol. 128, Pt. H, No. 1, Febr. 1981

SUPPLEMENT 2

Examples

S8 INTRODUCTION

- S8.1** The following examples are chosen to demonstrate further the method of evaluating the uncertainty of measurement. They supplement the examples presented in Supplement 1 to EA-4/02. The present collection of examples focuses on situations where there are one or two dominant terms in the uncertainty propagation or where the number of repeated measurements is small.
- S8.2** The examples are chosen to illustrate situations encountered in practice. It should be emphasised, however, that in practical applications there is no need to go through the mathematical derivations presented in these examples, in particular in the mathematical notes appended to some of the examples. Rather, the user is encouraged to employ the results of the theoretical presentations after having made himself acquainted with the conditions that have to be fulfilled. For instance, if it is ascertained, in a given situation, that the result of measurement has a rectangular distribution (as would be the case if there were only one term, rectangularly distributed, that needed to be considered in the propagation), one can immediately draw the conclusion that the coverage factor to be used to arrive at a coverage probability of 95 % is $k = 1,65$ (see S9.14).
- S8.3** One general conclusion that may be drawn from the uncertainty propagation is that in the case of only one dominant contribution the type of distribution of this contribution applies for the result of measurement as well. However, to evaluate the uncertainty of the result of measurement, the applicable sensitivity coefficient has to be employed, as usual.
- S8.4** It should be added that the situation where there is only one or a few dominant terms to the uncertainty of measurement is often met in connection with less complicated measuring instruments, where the dominant term often is due to the limited resolution of the instrument. Thus it may appear a paradox that the treatment of uncertainty of measurement for less complicated instruments, as shown by the examples of this Supplement, is more complicated than the treatment of the more straight-forward examples in Supplement 1. However, it should be kept in mind that the mathematical derivations, which may be felt as complications, are inserted for pedagogical reasons at places where they are needed instead of presenting them in the main document.
- S8.5** The examples are based on drafts prepared by EA Expert Groups. These drafts have been simplified and harmonised to make them transparent to the laboratory staff in all fields of calibration. It is thus hoped that this set of examples, like the preceding set published as Supplement 1 to EA-4/02, will contribute to a better understanding of the details of setting up the model of evaluation and to the harmonisation of the process of evaluating the uncertainty of measurement, independent of the field of calibration.
- S8.6** The contributions and values given in the examples are not intended to imply mandatory or preferred requirements. Laboratories should determine the uncertainty contributions on the basis of the model function they use in the evaluation of the particular calibration they perform and report the evaluated uncertainty of measurement on the calibration certificate they issue.

- S8.7** The presentation of the examples follows the common scheme presented and implemented in the first supplement to EA-4/02. For details the reader is referred to clause S1.4 of that document.
- S8.8** The uncertainty analysis of the examples is intended to represent the fundamentals of the specific measurement process and the method of evaluating the measurement result and the associated uncertainty. To keep the analysis transparent, also for those who are not experts in the relevant metrological field, a uniform method for the choice of the symbols of quantities has been followed, focused more on the physical background than on the current practice in different fields.
- S8.9** There are several recurrent quantities involved in all cases. One of them is the measurand, i.e., the quantity to be measured, another is the quantity presented by the working standard, which realises the local unit; with this quantity the measurand is compared. Besides these two quantities there are several others, in all cases, which take the role of additional local quantities or corrections.
- S8.10** Corrections describe the imperfect equality between a measurand and the result of a measurement. Some of the corrections are given by complete results of measurement, i.e., a measured value and its associated measurement uncertainty. For others the distribution of values is inferred from more or less complete knowledge of their nature. In most cases this will lead to an estimation of the limits for the unknown deviations.
- S8.11** In certain cases the quantity presented by a working standard is characterised by the nominal value of the standard. Thus, nominal values, which generally speaking characterise or identify calibration artefacts, often enter the uncertainty analysis.
- S8.12** To distinguish in the mathematical models of evaluation between these concepts, the examples have been designed to follow the notational rules given below. It is evident, however, that it is not possible to follow such rules strictly, because the practice concerning the use of symbols is different in different metrological fields.
- S8.13** The notation applied here distinguishes between main values, nominal values, correction values and values of limits:

Main values are measured or observed values that contribute an essential part to the value of a measurand. They are represented by lower-case letters in italics; they will be preceded by an upper-case Greek delta if the quantity represents a difference.

EXAMPLE:

t_{iX} - temperature indicated by a thermometer X to be calibrated. (Index i means indicated),

Δl - observed difference in the displacement of a measuring spindle.

Nominal values are assigned values of the realisation of a quantity by a standard or a measuring instrument. They are approximate values that give the main part of the realised value. They are represented by upper-case letters in italics.

EXAMPLE:

L - nominal length of a gauge block to be calibrated.

Correction values give small deviations from the main values that are known or have to be estimated. In most cases they are additive. They are represented by the symbol chosen for the quantity under consideration, preceded by a lower-case Greek delta.

EXAMPLE:

δm_D - possible deviation because of the drift of the value of a reference weight since its last calibration

δm_C - correction for eccentricity of load and magnetic effects in the calibration of a weight.

Values of limits are fixed, estimated values of possible variations of the unknown values of a quantity. They are represented by the symbol chosen for the quantity under consideration, preceded by an upper-case Greek delta.

EXAMPLE:

$\Delta\alpha_x$ - estimated half-width of the interval of possible deviations of a linear thermal resistivity coefficient given in a manufacturer's specification for a resistor to be calibrated.

The differentiation between different quantities of the same kind is affected by indices as shown in the examples. The internationally accepted notational rules for physical quantities have been followed: indices representing physical quantities are given in italics whereas indices that symbolise artefacts, instruments and so on are written in upright letters.

S8.14 Defined reference values are represented by a quantity symbol with the index zero.

EXAMPLE:

*p*₀ - reference pressure, e.g., of 1000 mbar.

S8.15 Ratios of quantities of the same kind (dimensionless ratios) are represented by lower-case letters in italics.

EXAMPLE:

$r = R_{iX} / R_{iN}$ - ratio of indicated resistance of an unknown resistor and a reference resistor (index i means indicated).

S8.16 If several indices are used, the sequence of indices is chosen in such a way that the index representing the most general concept is leftmost and the one representing the most specific concept is rightmost.

EXAMPLE:

V_{i1}, V_{i2} - voltage indicated by voltmeter '1' and voltmeter '2', respectively

S8.17 The examples in this second supplement to EA-4/02 are intended to be followed by others, illustrating different aspects encountered in connection with the calibration of measuring instruments. Examples may also be found in EA Guidance Documents dealing with the calibration of specific types of measuring instruments.

S9 CALIBRATION OF A HAND-HELD DIGITAL MULTIMETER AT 100 V DC

S9.1 As part of a general calibration, a hand-held digital multimeter (DMM) is calibrated at an input of 100 V DC using a multifunction calibrator as a working standard. The following measuring procedure is used:

- 1) The calibrator's output terminals are connected to the input terminals of the DMM using suitable measuring wires.
- 2) The calibrator is set to its 100 V setting and, after a suitable stabilising period, the DMM reading is noted.
- 3) The error of indication of the DMM is calculated using the DMM readings and the calibrator settings.

S9.2 It must be noted that the error of indication of the DMM which is obtained using this measuring procedure includes the effect of offset as well as deviations from linearity.

S9.3 The error of indication E_X of the DMM to be calibrated is obtained from

$$E_X = V_{iX} - V_S + \delta V_{iX} - \delta V_S \quad (\text{S9.1})$$

where

- V_{iX} - voltage, indicated by the DMM (index i means indication),
- V_S - voltage generated by the calibrator,
- δV_{iX} - correction of the indicated voltage due to the finite resolution of the DMM,

- δV_S - correction of the calibrator voltage due to
- (1) drift since its last calibration,
 - (2) deviations resulting from the combined effect of offset, non-linearity and differences in gain,
 - (3) deviations in the ambient temperature,
 - (4) deviations in mains power,
 - (5) loading effects resulting from the finite input resistance of the DMM to be calibrated.

S9.4 Because of the limited resolution of the indication of the DMM, no scatter in the indicated values is observed.

S9.5 DMM readings (V_{ix})

The DMM indicates the voltage 100,1 V at the calibrator setting 100 V. The DMM reading is assumed to be exact (see S9.4).

S9.6 Working standard (V_S)

The calibration certificate for the multifunction calibrator states that the voltage generated is the value indicated by the calibrator setting and that the associated expanded relative uncertainty of measurement is $W=0,000\ 02$ (coverage factor $k=2$), resulting in an expanded uncertainty of measurement associated with the 100 V setting of $U=0,002$ V (coverage factor $k=2$).

S9.7 Resolution of DMM to be calibrated (δV_{ix})

The least significant digit of the DMM display corresponds to 0,1 V. Each DMM reading has a correction due to the finite resolution of the display which is estimated to be 0,0 V with limits of ± 0.05 V (i.e., one half of the magnitude of the least significant digit).

S9.8 Other corrections (δV_S)

Because of the fact that individual figures are not available the uncertainty of measurement associated with the different sources is derived from the accuracy specification given by the manufacturer of the calibrator. These specifications state that the voltage generated by the calibrator coincides with the calibrator setting within $\pm(0,000\ 1 \times V_S + 1\ \text{mV})^1$ under the measuring conditions

- (1) the ambient temperature is within the range 18 °C to 23 °C
- (2) the mains voltage powering the calibrator is in the range 210 V to 250 V,
- (3) the resistive load at the calibrator's terminals is greater than 100 k Ω ,
- (4) the calibrator has been calibrated within the last year.

¹ A widely used method of presenting accuracy specification of measuring instruments in data sheets or manuals consists in giving the specification limits in terms of 'settings'. For the calibrator, the statement would be $\pm(0,01\%$ of setting + 1 mV). Even if this method is considered to be equivalent to the expression given above it is not used here because it may be misleading in many cases and because it does not represent an equation of physical quantities in the internationally accepted symbolic nomenclature.

Since these conditions of measurement are fulfilled and the calibration history of the calibrator shows that the manufacturer's specification may be relied upon, the correction to be applied to the voltage generated by the calibrator is assumed to be 0,0 V within $\pm 0,011$ V.

S9.9 Correlation

None of the input quantities is considered to be correlated to any significant extent.

S9.10 Uncertainty budget (E_X)

quantity x_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
V_{iX}	100,1 V	-	-	-	-
V_S	100,0 V	0,001 V	normal	-1,0	-0,001 V
δV_{iX}	0,0 V	0,029 V	rectangular	1,0	0,029 V
δV_S	0,0 V	0,0064 V	rectangular	-1,0	-0,0064 V
E_X	0,1 V				0,030 V

S9.11 Expanded uncertainty

The standard uncertainty of measurement associated with the result is clearly dominated by the effect of the finite resolution of the DMM. The final distribution is not normal but essentially rectangular. Therefore, the method of effective degrees of freedom described in Annex E of EA-4/02 is not applicable. The coverage factor appropriate for a rectangular distribution is calculated from the relation given in eq. (S9.8) in the mathematical note S9.14.

$$U = k \cdot u(E_X) = 1,65 \times 0,030 \text{ V} \cong 0,05 \text{ V}$$

S9.12 Reported result

The measured error of indication of the hand-held digital voltmeter at 100 V is (0,10 \pm 0,05) V.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 1,65$ which has been derived from the assumed rectangular probability distribution for a coverage probability of 95%.

S9.13 Additional remark

The method used for calculating the coverage factor is clearly related to the fact that the measurement uncertainty associated with the result is dominated by the effect of the finite resolution of the DMM. This will be true for the calibration of all low-resolution indicating instruments provided the finite resolution is the only dominant source in the uncertainty budget.

S9.14 Mathematical note

If the situation of measurement is such that one of the uncertainty contributions in the budget can be identified as a dominant term, for instance the term with index 1, the standard uncertainty to be associated with the measurement result y can be written as

$$u(y) = \sqrt{u_1^2(y) + u_R^2(y)}. \quad (\text{S9.2})$$

Here is

$$u_R(y) = \sqrt{\sum_{i=2}^N u_i^2(y)} \quad (\text{S9.3})$$

denotes the total uncertainty contribution of the non-dominant terms. As long as the ratio of the total uncertainty contribution $u_R(y)$ of the non-dominant terms to the uncertainty contribution $u_1(y)$ of the dominant term is not larger than 0,3, eq. (S9.2) may be approximated by

$$u(y) \cong u_1(y) \cdot \left[1 + \frac{1}{2} \left(\frac{u_R(y)}{u_1(y)} \right)^2 \right]. \quad (\text{S9.4})$$

The relative error of approximation is smaller than 1×10^{-3} . The maximum relative change in the standard uncertainty resulting from the factor within the brackets in eq. (S9.4) is not larger than 5%. This value is within the accepted tolerance for mathematical rounding of uncertainty values.

Under these assumptions the distribution of values that could reasonably be attributed to the measurand is essentially identical with the distribution resulting from the known dominant contribution. From this distribution density $\varphi(y)$ the coverage probability p may be determined for any value of the expanded measurement uncertainty U by the integral relation

$$p(U) = \int_{y-U}^{y+U} \varphi(y') dy'. \quad (\text{S9.5})$$

Inverting this relation for a given coverage probability results in the relation between the expanded measurement uncertainty and the coverage probability $U = U(p)$ for the given distribution density $\varphi(y)$. Using this relation, the coverage factor may finally be expressed as

$$k(p) = \frac{U(p)}{u(y)}. \quad (\text{S9.6})$$

In the case of the hand-held digital voltmeter the dominant uncertainty contribution resulting from the finite resolution of the indication is $u_{\delta V_X}(E_X) = 0,029 \text{ V}$ whereas the total uncertainty contribution of the non-dominant terms is $u_R(E_X) = 0,0065 \text{ V}$. The relevant ratio is $u_R(E_X)/u_{\delta V_X}(E_X) = 0,22$. Thus, the resulting distribution of values that can reasonably be attributed as errors of indications is essentially rectangular. The coverage probability for a rectangular distribution is linearly related to the expanded measurement uncertainty (a being the half-width of the rectangular distribution)

$$p = \frac{U}{a} . \quad (\text{S9.7})$$

Solving this relation for the expanded measurement uncertainty U and inserting the result together with the expression of the standard measurement uncertainty related to a rectangular distribution as given by eq. (3.8) of EA-4/02 finally gives the relation

$$k(p) = p\sqrt{3} . \quad (\text{S9.8})$$

For a coverage probability $p = 95 \%$ applicable in the EA, the relevant coverage factor is thus $k = 1,65$.

S10 CALIBRATION OF A VERNIER CALLIPER

S10.1 A vernier calliper made of steel is calibrated against grade I gauge blocks of steel used as working standards. The measurement range of the calliper is 150 mm. The reading interval of the calliper is 0,05 mm (the main scale interval is 1 mm and the vernier scale interval 1/20 mm). Several gauge blocks with nominal lengths in the range (0,5 – 150) mm are used in the calibration. They are selected in such a way that the measurement points are spaced at nearly equal distances (e.g. at 0 mm, 50 mm, 100 mm, 150 mm) but give different values on the vernier scale (e.g. 0,0 mm, 0,3 mm, 0,6 mm, 0,9 mm). The example concerns the 150 mm calibration point for measurement of external dimensions. Before calibration several checks of the condition of the calliper are made. These include dependence of the result of measurement on the distance of the measured item from the beam (Abbe error), quality of the measuring faces of the jaws (flatness, parallelism, squareness), and function of the locking mechanism.

S10.2 The error of indication E_X of the calliper at the reference temperature $t_0 = 20^\circ\text{C}$ is obtained from the relation:

$$E_X = l_{iX} - l_S + L_S \cdot \bar{\alpha} \cdot \Delta t + \delta l_{iX} + \delta l_M \quad (\text{S10.1})$$

where:

l_{iX}	-	indication of the calliper,
l_S	-	length of the actual gauge block,
L_S	-	nominal length of the actual gauge block,
$\bar{\alpha}$	-	average thermal expansion coefficient of the calliper and the gauge block,
Δt	-	difference in temperature between the calliper and the gauge block,
δl_{iX}	-	correction due to the finite resolution of the calliper,
δl_M	-	correction due to mechanical effects, such as applied measurement force, Abbe errors, flatness and parallelism errors of the measurement faces.

S10.3 Working standards (l_S, L_S)

The lengths of the reference gauge blocks used as working standards, together with their associated expanded uncertainty of measurement, are given in the calibration certificate. This certificate confirms that the gauge blocks comply with the requirements for grade I gauge blocks according to ISO 3650, i.e. that the central length of the gauge blocks coincides within $\pm 0,8 \mu\text{m}$ with the nominal length. For the actual lengths of the gauge blocks their nominal lengths are used without correction, taking the tolerance limits as the upper and lower limits of the interval of variability.

S10.4 Temperature ($\Delta t, \bar{\alpha}$)

After an adequate stabilisation time, the temperatures of the calliper and the gauge block are equal within $\pm 2 \text{ }^\circ\text{C}$. The average thermal expansion coefficient is $11,5 \cdot 10^{-6} \text{ }^\circ\text{C}^{-1}$. (The uncertainty in the average thermal expansion coefficient and in the difference of the thermal expansion coefficients has not been taken into account; its influence is considered negligible for the present case. Cf. EA-4/02-S1, example S4.)

S10.5 Resolution of the calliper (δl_{iX})

The scale interval of the vernier scale is 0,05 mm. Thus variations due to the finite resolution are estimated to have rectangular limits of $\pm 25 \mu\text{m}$.

S10.6 Mechanical effects (δl_M)

These effects include the applied measurement force, the Abbe error and the play between the beam and the sliding jaw. Additional effects may be caused by the fact that the measuring faces of the jaws are not exactly flat, not parallel to each other and not perpendicular to the beam. To minimise effort, only the range of the total variation, equal to $\pm 50 \mu\text{m}$ is considered.

S10.7 Correlation

None of the input quantities is considered to be correlated to any significant extent.

S10.8 Measurements (l_{iX})

The measurement is repeated several times without detecting any scatter in the observations. Thus uncertainty due to limited repeatability does not give a contribution. The result of measurement for the 150 mm gauge block is 150,10 mm.

S10.9 Uncertainty budget (δl_x)

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
l_{iX}	150,10 mm	-	-	-	-
l_S	150,00 mm	0,46 μm	rectangular	-1,0	-0,46 μm
Δt	0	1,15 K	rectangular	1,7 $\mu\text{m}\cdot\text{K}^{-1}$	2,0 μm
δl_{iX}	0	14 μm	rectangular	1,0	14 μm
δl_M	0	29 μm	rectangular	1,0	29 μm
E_X	0,10 mm				32 μm

S10.10 Expanded uncertainty

The uncertainty of measurement associated with the result is clearly dominated by the combined effect of the measurement force and the finite resolution of the vernier. The final distribution is not normal but essentially trapezoidal with a ratio $\beta = 0,33$ of the half-width of the plateau region to the half-width of the variability interval. Therefore the method of effective degrees of freedom described in EA-4/02, Annex E is not applicable. The coverage factor $k = 1,83$ appropriate for this trapezoidal distribution of values is calculated from eq. (S10.10) of the mathematical note S10.13. Thus

$$U = k \cdot u(E_X) = 1,83 \cdot 0,032 \text{ mm} \cong 0,06 \text{ mm}$$

S10.11 Reported result

At 150 mm the error of indication of the calliper is $(0,10 \pm 0,06)$ mm.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 1,83$ which has been derived from the assumed trapezoidal probability distribution for a coverage probability of 95 %.

S10.12 Additional remark

The method used for calculating the coverage factor is clearly related to the fact that uncertainty of measurement associated with the result is dominated by two influences: the mechanical effects and the finite resolution of the vernier scale. Thus the assumption of a normal distribution for the output quantity is not justified and the conditions of EA-4/02, paragraph 5.6 apply. In the sense that probabilities and probability densities in practice may only be determined to within 3 %– 5 %, the distribution is essentially trapezoidal, obtained by convolution of the two rectangular distributions associated with the dominant contributions. The half-widths of the base and the top of the resulting symmetrical trapezoid are 75 μm and 25 μm , respectively. 95 % of the area of the trapezoid is encompassed by an interval $\pm 60 \mu\text{m}$ around its symmetry axis, corresponding to $k = 1,83$.

S10.13 Mathematical note

If the situation of measurement is such that two of the uncertainty contributions in the budget can be identified as dominant terms, the method presented in S9.14 can be applied when the two dominant contributions, for instance the terms with indices 1 and 2, are combined into one dominant term. The standard uncertainty to be associated with the measurement result y may be written in this case as

$$u(y) = \sqrt{u_0^2(y) + u_R^2(y)} \quad (\text{S10.2})$$

where

$$u_0(y) = \sqrt{u_1^2(y) + u_2^2(y)} \quad (\text{S10.3})$$

denotes the combined contribution of the two dominant terms and

$$u_R(y) = \sqrt{\sum_{i=3}^N u_i^2(y)} \quad (\text{S10.4})$$

the total uncertainty contribution of the remaining non-dominant terms. If the two dominant contributions arise from rectangular distributions of values with half-widths a_1 and a_2 , the distribution resulting from convolving them is a symmetrical trapezoidal distribution




Fig. 1: Unified symmetrical trapezoidal probability distribution with the value $\beta=0,33$ of the edge parameter, resulting from the convolution of two rectangular distributions.

with half-widths

$$a = a_1 + a_2 \text{ and } b = |a_1 - a_2| \quad (\text{S10.5})$$

of the base and the top, respectively (see example in Fig. 1). The distribution may be conveniently expressed in the unified form

$$\varphi(y) = \frac{1}{a(1+\beta)} \times \begin{cases} 1 & |y| < \beta \cdot a \\ \frac{1}{1-\beta} \left(1 - \frac{|y|}{a}\right) & \beta \cdot a \leq |y| \leq a \\ 0 & a < |y| \end{cases} \quad (\text{S10.6})$$

with the edge parameter

$$\beta = \frac{b}{a} = \frac{|a_1 - a_2|}{a_1 + a_2} \quad (\text{S10.7})$$

The square of the standard measurement uncertainty deduced from the trapezoidal distribution of eq. (S10.6) is

$$u^2(y) = \frac{a^2}{6} (1 + \beta^2) . \quad (\text{S10.8})$$

Using the distribution of eq. (S10.6) the dependence of the coverage factor on the coverage probability is derived according to the method sketched in S9.14

$$k(p) = \frac{1}{\sqrt{\frac{1+\beta^2}{6}}} \times \begin{cases} \frac{p(1+\beta)}{2} & \frac{p}{2-p} < \beta \\ 1 - \sqrt{(1-p)(1-\beta^2)} & \beta \leq \frac{p}{2-p} \end{cases} . \quad (\text{S10.9})$$

Fig. 2 shows the dependence of the coverage factor k on the value of the edge parameter β for a coverage probability of 95 %.




Fig. 2: Dependence of the coverage factor k on the value of the edge parameter β of a trapezoidal distribution for a coverage probability of 95 %.

The coverage factor for a coverage probability of 95 % appropriate to a trapezoidal distribution with an edge parameter of $\beta < 0,95$ is calculated from the relation

$$k = \frac{1 - \sqrt{(1-p)(1-\beta^2)}}{\sqrt{\frac{1+\beta^2}{6}}} \quad (\text{S10.10})$$

S11 CALIBRATION OF A TEMPERATURE BLOCK CALIBRATOR AT A TEMPERATURE OF 180 °C²

S11.1 As part of a calibration, the temperature that has to be assigned to the calibration bore of a temperature block calibrator, is measured. This is done when the indication of the built-in temperature indicator has stabilised at 180,0 °C. The temperature of the calibration bore is determined by an inserted platinum resistance thermometer, used as a working standard, by measuring the electrical resistance of the thermometer by an ac resistance bridge. The temperature t_X , that has to be assigned as the temperature of the bore when the reading of the built-in temperature indicator is 180,0 °C, is given by:

$$t_X = t_S + \delta t_S + \delta t_D - \delta t_{iX} + \delta t_R + \delta t_A + \delta t_H + \delta t_V \quad (\text{S11.1})$$

where:

- t_S - temperature of the working standard derived from the ac resistance measurement,
- δt_S - temperature correction due to the ac resistance measurement,
- δt_D - temperature correction due to drift in the value of the working standard since its last calibration,
- δt_{iX} - temperature correction due to the settability limitations of the block temperature calibrator,
- δt_R - temperature correction due to the radial temperature difference between the built-in thermometer and the working standard,
- δt_A - temperature correction due to the axial inhomogeneity of temperature in the measuring bore,
- δt_H - temperature correction due to hysteresis in the increasing and decreasing branches of the measuring cycle,
- δt_V - temperature variation within the time of measurement.

Temperature corrections due to stem conduction are not considered, since the platinum resistance thermometer used as working standard has an outer diameter $d \leq 6$ mm. Prior investigations have shown that stem conduction effects can be neglected in this case.

² This process is basic for calibrations in different metrological fields and, therefore, of general interest. The example further demonstrates that there are two equivalent ways to tackle this problem: the direct assignment of a value to the indication of the instrument and the association of a correction to the indication, usually called the error of indication.

S11.2 Working standard (t_S)

The calibration certificate of the resistance thermometer used as working standard gives the relationship between resistance and temperature. The measured resistance value corresponds to a temperature of 180,1 °C, with an associated expanded uncertainty of measurement $U = 30$ mK (coverage factor $k = 2$).

S11.3 Determination of the temperature by resistance measurement (δt_S)

The temperature of the resistance thermometer used as working standard is determined as 180,1 °C. The standard measurement uncertainty associated with the resistance measurement converted to temperature corresponds to $u(\delta t_S) = 10$ mK.

S11.4 Drift of the temperature of the working standard (δt_D)

From general experience with platinum resistance thermometers of the type used as working standard in the measurement, the change of temperature due to resistance ageing since the last calibration of the standard is estimated to be within the limits ± 40 mK.

S11.5 Settability of the block temperature calibrator (δt_X)

The built-in controlling thermometer of the block temperature calibrator has a scale interval of 0,1 K. This gives temperature resolution limits of ± 50 mK within which the thermodynamic state of the temperature block can be uniquely set.

Note: If the indication of the built-in temperature indicator is not given in units of temperature the resolution limits must be converted into equivalent temperature values by multiplying the indication with the relevant instrument constant.

S11.6 Radial inhomogeneity of temperature (δt_R)

The radial temperature difference between the measuring bore and the built-in thermometer has been estimated to be within ± 100 mK.

S11.7 Axial inhomogeneity of temperature (δt_A)

The temperature deviations due to axial inhomogeneity of temperature in the calibration bore have been estimated from readings for different immersion depths to be within ± 250 mK.

S11.8 Hysteresis effects (δt_H)

From readings of the reference thermometer during measurement cycles of increasing and decreasing temperature, the temperature deviation of the calibration bore due to hysteresis effect has been estimated to be within ± 50 mK.

S11.9 Temperature instability (δt_V)

Temperature variations due to temperature instability during the measuring cycle of 30 min are estimated to be within ± 30 mK.

S11.10 Correlations

None of the input quantities is considered to be correlated to any significant extent.

S11.11 Repeated observations

Due to the finite resolution of the indication of the built-in thermometer no scatter in the indicated values has been observed and taken into account.

S11.12 Uncertainty budget (t_X)

Quantity X_i	Estimate x_i	Standard uncertainty $u(x_i)$	Probability distribution	Sensitivity coefficient c_i	Uncertainty contribution $u_i(y)$
t_S	180,1 °C	15 mK	normal	1,0	15 mK
δt_S	0,0 °C	10 mK	normal	1,0	10 mK
δt_D	0,0 °C	23 mK	rectangular	1,0	23 mK
δt_{iX}	0,0 °C	29 mK	rectangular	-1,0	-29 mK
δt_R	0,0 °C	58 mK	rectangular	1,0	58 mK
δt_A	0,0 °C	144 mK	rectangular	1,0	144 mK
δt_H	0,0 °C	29 mK	rectangular	1,0	29 mK
δt_V	0,0 °C	17 mK	rectangular	1,0	17 mK
t_X	180,1 °C				164 mK

S11.13 Expanded uncertainty

The standard uncertainty of measurement associated with the result is clearly dominated by the effect of the unknown temperature correction due to the axial temperature inhomogeneity in the measuring bore and the radial temperature difference between the built-in thermometer and the working standard. The final distribution is not normal but essentially trapezoidal. According to S10.13, the coverage factor corresponding to the edge parameter $\beta = 0,43$ is $k = 1,81$.

$$U = k \cdot u(t_X) = 1,81 \times 164 \text{ mK} \cong 0,3 \text{ K}$$

S11.14 Reported result

The temperature boring of the calibration bore that has to be assigned to an indication of the built-in controlling thermometer of 180,0 °C is (180,1 ± 0,3) °C.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 1,81$ which has been derived from the assumed trapezoidal probability distribution for a coverage probability of 95 %.

S11.15 Mathematical note concerning the model

Some metrologists are confused that the indication of the controlling thermometer does not appear explicitly in the model function of eq. (S11.1). To fit their needs, the problem can alternatively be formulated with the error of indication

$$E_X = t_X - t_i \quad (\text{S11.2})$$

of the built-in temperature indicator

$$E_X = t_S - t_i + \delta t_S + \delta t_D - \delta t_{iX} + \delta t_R + \delta t_A + \delta t_H + \delta t_V \quad (\text{S11.3})$$

The indicated value t_i is a nominal value. Its effect is to shift the scale of the measurand. It does, however, not contribute to the uncertainty of measurement associated with the error of indication

$$u(E_X) = u(t_X) \quad (\text{S11.4})$$

The model function of eq. (S11.1) can be regained from eq. (S11.3) using the definition of the error of indication in eq. (S11.2).

This note shows that there is not necessarily only one unique way to choose the model of evaluation of measurement. The metrologist keeps it in his hands to choose the model that suits his habits and his approach to the problem. Model functions that can be transformed mathematically from one into the other represent the same measurement process. For cases where a continuous scale of indication is involved, as in the calibration of the temperature block under consideration, model functions that are connected by linear scale transformations may serve as equivalent expressions of the measurement problem.

S12 CALIBRATION OF A HOUSEHOLD WATER METER

S12.1 The calibration of a water meter involves the determination of the relative error of indication within the applicable flow range of the meter. The measurement is made using a test rig that supplies necessary water flow with a pressure of approximately 500 kPa, a value typical for municipal tap water systems. The water is received in an open collecting tank that has been calibrated and determines the reference volume of the water. It is empty but wetted at the beginning of the measurement. The collecting tank has a narrow neck with an attached scale by which the filling level can be detected. The meter to be calibrated is connected between these tanks. It has a mechanical counter with pointers. The measurement is done at a flow rate of 2500 l/h with standing start-and-stop which means that the flow rate is zero both at the beginning and the end of the measurement. The indication of the meter is recorded at the beginning and at the end of the measurement. The level is recorded in the collecting tank at the end of the measurement. The temperature and pressure of the water at the meter, and the temperature of the water in the collecting tank, are recorded as well.

S12.2 The relative error of indication e_X in a single run is defined as

$$e_X = \frac{\Delta V_{iX} + \delta V_{iX2} - \delta V_{iX1}}{V_X} - 1 \quad (\text{S12.1})$$

with

$$V_X = (V_{iS} + \delta V_{iS})(1 + \alpha_S(t_S - t_0))(1 + \alpha_W(t_X - t_S))(1 - \kappa_W(p_X - p_S)) \quad (\text{S12.2})$$

where:

$\Delta V_{iX} = V_{iX2} - V_{iX1}$	- difference in meter indications,
V_{iX1}, V_{iX2}	- indication of the meter at the beginning of the measurement and at the end of the measurement,
$\delta V_{iX1}, \delta V_{iX2}$	- corrections due to the finite resolution of the meter indication,
V_X	- volume that passed the meter during the measurement under the prevailing conditions, i.e., pressure p_X and temperature t_X , at the inlet of the meter,
V_{iS}	- volume indicated at the neck scale of the collecting tank at the end of the measurement,
δV_{iS}	- correction of the volume indicated at the neck scale of the collecting tank due to the finite resolution of the scale,
α_S	- cubic thermal expansion coefficient of the material of the collecting tank,
t_S	- temperature of the collecting tank,
t_0	- reference temperature at which the collecting tank has been calibrated,
α_W	- cubic thermal expansion coefficient of water,
t_X	- temperature of the water at the inlet of the meter,
κ_W	- compressibility of water,
p_S	- pressure in the collecting tank (it is zero if excess pressure is considered)
p_X	- pressure of the water at the inlet of the meter.

S12.3 Collecting tank (V_{iS}, t_0)

The calibration certificate states that the neck scale indicates the volume of 200 l at the reference temperature $t_0 = 20^\circ\text{C}$ with an associated relative expanded measurement uncertainty of 0,1 % ($k = 2$). The expanded measurement uncertainty associated with the value is 0,2 l ($k = 2$).

S12.4 Resolution of the collecting tank scale (δV_{iS})

The water level of the collecting tank can be determined to within ± 1 mm. With the scale factor of the tank of 0,02 l/mm the maximum deviation of the volume of water in the collecting tank from the observed indicated value is estimated to be within $\pm 0,02$ l.

S12.5 Temperature of the water and the collecting tank (α_S, t_S)

The temperature of the water in the collecting tank is determined to be 15°C within ± 2 K. The stated limits cover all possible sources of uncertainty, such as calibration

of temperature sensors, resolution in reading and temperature gradients in the tank. The cubic thermal expansion coefficient of the tank material (steel) is taken from a material handbook to be a constant equal to $\alpha_s = 51 \cdot 10^{-6} \text{ K}^{-1}$ in the temperature interval considered. Since there is no uncertainty statement accompanying this value it is assumed to be known to within its least significant digit. Unknown deviations are considered to be within the rounding limits of $\pm 0,5 \cdot 10^{-6} \text{ K}^{-1}$.

S12.6 Temperature of the water at the meter (α_w, t_x)

The temperature of the water at the inlet of the meter is determined to be 16 °C within $\pm 2 \text{ K}$. The stated limits cover all possible sources of uncertainty, such as contributions from calibration of sensors, resolution in reading and small temperature changes during one measurement run. The cubic expansion coefficient of water is taken from a material handbook to be a constant equal to $\alpha_w = 0,15 \cdot 10^{-3} \text{ K}^{-1}$ in the temperature interval considered. Since there is no uncertainty statement accompanying this value it is assumed to be known to within its least significant digit. Unknown deviations are considered to be within the rounding limits of $\pm 0,005 \cdot 10^{-3} \text{ K}^{-1}$.

S12.7 Pressure difference of the water between the meter and the tank (κ_w, p_s, p_x)

The excess pressure of the water supplied to the inlet of the meter is 500 kPa with relative deviations not larger than $\pm 10 \%$. On its way from the inlet to the collecting tank, the water expands to excess pressure 0 kPa (atmospheric pressure condition). The compressibility of water is taken from a material handbook to be a constant equal to $\kappa_w = 0,46 \cdot 10^{-6} \text{ kPa}^{-1}$ in the temperature interval considered. Since there is no uncertainty statement accompanying this value, it is assumed to be known to within its least significant digit. Unknown deviations are considered to be within the rounding limits of $\pm 0,005 \cdot 10^{-6} \text{ kPa}^{-1}$.

S12.8 Correlation

None of the input quantities is considered to be correlated to any significant extent.

S12.9 Uncertainty budget (V_X)

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
V_{iS}	200,00 l	0,10 l	normal	1,0	0,10 l
δV_{iS}	0,0 l	0,0115 l	rectangular	1,0	0,0115 l
α_S	$51 \cdot 10^{-6} \text{ K}^{-1}$	$0,29 \cdot 10^{-6} \text{ K}^{-1}$	rectangular	-1000 l·K	$-0,29 \cdot 10^{-3} \text{ l}$
t_S	15°C	1,15 K	rectangular	$-0,0198 \text{ l} \cdot \text{K}^{-1}$	-0,0228 l
α_W	$0,15 \cdot 10^{-3} \text{ K}^{-1}$	$2,9 \cdot 10^{-6} \text{ K}^{-1}$	rectangular	200 l·K	$0,58 \cdot 10^{-3} \text{ l}$
t_X	16°C	1,15 K	rectangular	$0,0300 \text{ l} \cdot \text{K}^{-1}$	0,0346 l
κ_W	$0,46 \cdot 10^{-6} \text{ kPa}^{-1}$	$2,9 \cdot 10^{-9} \text{ kPa}^{-1}$	rectangular	$-1,00 \cdot 10^5 \text{ l} \cdot \text{kPa}^{-1}$	$-0,29 \cdot 10^{-3} \text{ l}$
p_X	500 kPa	29 kPa	rectangular	$-9,2 \cdot 10^{-5} \text{ l} \cdot \text{kPa}^{-1}$	-0,0027 l
p_S	0,0 Pa	-	-	-	-
V_X	199,93 l				0,109 l

The standard uncertainty of measurement associated with the result is clearly dominated by the volume indication at the neck scale of the collecting tank. The final distribution is essentially normal. This must be kept in mind in the further processing of the uncertainty evaluation.

S12.10 Indication of the meter (ΔV_{iX} , δV_{iX1} , δV_{iX2})

The water meter to be calibrated has a resolution of 0,2 l resulting in the limits $\pm 0,1$ l in both readings for the maximum deviations resulting from the meter resolution.

S12.11 Uncertainty budget (e_X)

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
ΔV_{iX}	200,0 l	-	nominal	-	-
δV_{iX1}	0,0 l	0,058 l	rectangular	$-5,0 \cdot 10^{-3}$	$-0,29 \cdot 10^{-3} \text{ l}$
δV_{iX2}	0,0 l	0,058 l	rectangular	$5,0 \cdot 10^{-3}$	$0,29 \cdot 10^{-3} \text{ l}$
V_X	199,93 l	0,109 l	normal	$-5,0 \cdot 10^{-3}$	$-0,55 \cdot 10^{-3} \text{ l}$
e_X	0,000 3				$0,68 \cdot 10^{-3}$

S12.12 Repeatability of the meter

The relative error of indication of the water meter to be calibrated, determined at the same flow rate of 2500 l/h, shows considerable scatter. For that reason, the relative error of indication is determined three times. The results of these three runs are treated as independent observations e_{Xj} in the model that determines the average error of indication e_{Xav} :

$$e_{Xav} = e_X + I/e_X$$

Hence

$$I/e_X = e_{Xav} - e_X$$

(S12.3)

where:

- e_X - relative error of indication of a single run,
- δe_X - correction of the relative error of indication obtained in the different runs due to the lack of repeatability of the meter.

S12.13 Measurements (e_X)

No.	observed relative error of indication	relative error due to lack of repeatability
1	0,000 3	0,0007
2	0,000 5	0,0005
3	0.002 2	-0,0012

average error of indication e_{Xav} is the arithmetic mean of the three results:

$$e_{Xav} = 0,001$$

average error due to lack of repeatability: $\bar{\delta}_{e_X} = 0$

experimental standard deviation: $s(I/e_X) = 0,001$

standard uncertainty: $u(\delta_{e_X}) = s(\bar{\delta}_{e_X}) = \frac{s(\delta_{e_X})}{\sqrt{3}} = 0,000 60$

S12.14 Uncertainty budget ($e_{X_{av}}$)

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	degrees of freedom ν_{eff}	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
δe_x	0	$0,60 \cdot 10^{-3}$	2	normal	1,0	$0,60 \cdot 10^{-3}$
e_x	0,001	$0,68 \cdot 10^{-3}$	∞	normal	1,0	$0,68 \cdot 10^{-3}$
$e_{X_{av}}$	0,001		10			$0,91 \cdot 10^{-3}$

S12.15 Expanded uncertainty

Because of the small number of effective degrees of freedom of the standard uncertainty associated with the mean relative error of indication the standard coverage factor has to be modified according to table E1

$$U = k \cdot u(e_{X_{av}}) = 2,28 \times 0,91 \cdot 10^{-3} \cong 2 \cdot 10^{-3}$$

S12.16 Reported result

The average relative error of indication of the water meter determined at a flow rate of 2500 l/h is $0,001 \pm 0,002$.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S13 CALIBRATION OF A RING GAUGE WITH A NOMINAL DIAMETER OF 90 MM

S13.1 A steel ring gauge of $D_x = 90$ mm nominal inner diameter is calibrated applying the procedure introduced in EURAMET cg-6 Version 2.0 (03/2011)" (previously EAL-G29:1997). A length comparator of the Abbe type and a steel setting ring, whose nominal inner diameter ($D_s = 40$ mm) differs significantly from that of the ring to be calibrated, are employed. In this case the length comparator and the steel setting ring both take the role of working standards. The rings are gently clamped sequentially on a 4-degrees of freedom table, which includes all position elements for aligning the test pieces. The rings are contacted at several points diametrically apart by two C-shaped arms, fixed on the steady and the measuring spindle, respectively. The C-shaped arms are supplied with spherical contact tips. The measuring force is generated by a tension weight ensuring a constant force of nominally 1,5 N over the whole measuring range. The measuring spindle is rigidly connected with the gauge head of a steel line scale of resolution $0,1 \mu\text{m}$. The line scale of the comparator has been verified periodically to meet the manufacturer's specification of maximum permissible error.

The ambient temperature is monitored in order to maintain the environmental conditions stated by calibration procedure. The temperature in the comparator working volume is maintained at $20\text{ °C} \pm 0,5\text{ K}$. Care is taken to ensure that the rings and the line scale (ruler) maintain the monitored temperature throughout the calibration.

S13.2 The diameter d_X of the ring to be calibrated at the reference temperature $t_0 = 20\text{ °C}$ is obtained from the relationship:

$$d_X = d_S + \Delta l + \delta l_i + \delta l_T + \delta l_P + \delta l_E + \delta l_A \quad (\text{S13.1})$$

where:

- d_S - diameter of the reference setting ring at the reference temperature,
- Δl - observed difference in displacement of the measuring spindle when the contact tips touch the inner surface of the rings at two diametrically apart points,
- δl_i - correction for the errors of indication of the comparator,
- δl_T - correction due to the temperature effects of the ring to be calibrated, the reference setting ring and the comparator line scale,
- δl_P - correction due to coaxial misalignment of the probes with respect to the measuring line,
- δl_E - correction due to the difference in elastic deformations of the ring to be calibrated and the reference setting ring,
- δl_A - correction due to the difference of the Abbe errors of the comparator when the diameters of the ring to be calibrated and the reference setting ring are measured.

S13.3 Working standard (d_S)

The inner diameter of the setting ring used as the working standard together with the associated expanded uncertainty of measurement is given in the calibration certificate as $40,0007\text{ mm} \pm 0,2\text{ }\mu\text{m}$ (coverage factor $k = 2$).

S13.4 Comparator (δl_i)

The corrections for the errors of indication of the line scale (ruler) were determined by the manufacturer and prestored electronically. Any residuals are within the manufacturers' specification of $\pm(0,3\text{ }\mu\text{m} + 1,5 \cdot 10^{-6} \cdot l_i)$ with l_i being the indicated length. The specifications are ascertained by periodical verifications. For the actual length difference $D_X - D_S = 50\text{ mm}$ unknown residuals are estimated to be within $\pm(0,375)\text{ }\mu\text{m}$.

S13.5 Temperature corrections (δl_T)

Throughout the measurement care is taken to ensure that the ring to be calibrated, the setting ring and the comparator scale maintain the monitored temperature. From previous measurements and general experience with the measurement system it can be ascertained that the deviations of temperatures of the ring to be calibrated, the

setting ring and the comparator scale from ambient temperature stay within $\pm 0,2$ K. The ambient temperature of the measuring room, however, is estimated to be within $\pm 0,5$ K. The knowledge on the measurement, therefore, is best described by the deviation of the ambient temperature from the reference temperature and the deviations of the temperatures of the ring to be calibrated, the setting ring and the comparator scale (ruler) from the ambient temperature. The correction δl_T due to temperature influences is determined from the model:

$$\begin{aligned} \delta l_T = & (D_S \cdot (\alpha_S - \alpha_R) - D_X \cdot (\alpha_X - \alpha_R)) \cdot \Delta t_A \\ & + D_S \cdot \alpha_S \cdot \delta t_S - D_X \cdot \alpha_X \cdot \delta t_X - (D_S - D_X) \cdot \alpha_R \cdot \delta t_R \end{aligned} \quad (\text{S13.2})$$

where:

- D_X, D_S - nominal diameters of the ring to be calibrated and the reference setting ring,
- $\alpha_X, \alpha_S, \alpha_R$ - linear thermal expansion coefficients of the ring to be calibrated, the reference setting ring and the comparator line scale (ruler),
- $\Delta t_A = t_A - t_0$ - deviations of the ambient temperature of the measuring room from the reference temperature $t_0 = 20^\circ\text{C}$,
- $\delta t_X, \delta t_S, \delta t_R$ - deviations of the temperature of the ring to be calibrated, the reference setting ring and the comparator line scale (ruler) from ambient temperature

Since the expectations of the four temperature differences entering eq. (S13.2) are zero, the usual linearized version will not include effects of the measurement uncertainty associated with the values of the three linear thermal expansion coefficients. As depicted in section S4.13 the non-linear version has to be used to determine the standard uncertainty associated with the four product terms:

$$\begin{aligned} \delta l_{TA} &= (D_S \cdot (\alpha_S - \alpha_R) - D_X \cdot (\alpha_X - \alpha_R)) \cdot \Delta t_A \\ \delta l_{TS} &= D_S \cdot \alpha_S \cdot \delta t_S \\ \delta l_{TX} &= D_X \cdot \alpha_X \cdot \delta t_X \\ \delta l_{TR} &= (D_S - D_X) \cdot \alpha_R \cdot \delta t_R \end{aligned} \quad (\text{S13.3})$$

Based on the calibration certificate of the setting ring, on the manufacturer's data for the ring to be calibrated and the comparator scale, the linear thermal expansion coefficients are assumed to be within the interval $(11,5 \pm 1,0) 10^{-6} \text{ K}^{-1}$. Using this value and the limits of temperature variation stated at the beginning, the standard uncertainties associated with the four product terms are $u(\delta l_{TA}) = 0,018 \mu\text{m}$, $u(\delta l_{TS}) = 0,053 \mu\text{m}$, $u(\delta l_{TX}) = 0,12 \mu\text{m}$ and $u(\delta l_{TR}) = 0,066 \mu\text{m}$. The standard uncertainty associated with the combined temperature corrections is derived from these values with the use of the following uncertainty sub-budget:

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
δl_{TA}	0,0 μm	0,018 μm	-	1,0	0,018 μm
δl_{TS}	0,0 μm	0,053 μm	-	1,0	0,053 μm
δl_{TX}	0,0 μm	0,12 μm	-	1,0	0,12 μm
δl_{TR}	0,0 μm	0,066 μm	-	1,0	0,066 μm
δl_T	0,0 μm				0,15 μm

S13.6 Coaxiality correction (δl_p)

The deviation from coaxiality of the two spherical probes and the measuring line is assumed to be within $\pm 20 \mu\text{m}$. Using the equations stated in the mathematical note (S13.13) the correction due to possible non-coaxiality and the associated standard uncertainty of measurement is given by

$$\delta l_p = 2 \cdot \left(\frac{1}{D_x} - \frac{1}{D_s} \right) \cdot u^2(\delta c) \quad (\text{S13.4})$$

$$u^2(\delta l_p) = \frac{16}{5} \cdot \left(\frac{1}{D_x^2} + \frac{1}{D_s^2} \right) \cdot u^4(\delta c) \quad (\text{S13.5})$$

Here δc is the small distance of the measured cord from the centre of the ring. The values resulting for the correction and the associated standard measurement uncertainty are $\delta l_p \cong -0,004 \mu\text{m}$ and $u(\delta l_p) \cong 0,0065 \mu\text{m}$. As can be seen from the uncertainty budget (S13.10), these values are two orders of magnitude smaller than the remaining uncertainty contributions so that their influence need not be taken into account under the current measurement conditions.

S13.7 Elastic deformation correction (δl_E)

The elastic deformation of the ring to be calibrated or the reference setting ring are not determined during the current measurement. From previous experience, however, the effects resulting from elastic deformations are estimated to be within $\pm 0,03 \mu\text{m}$.

S13.8 Abbe error correction (δl_A)

The actual values of the Abbe errors of the comparator are not determined during the current measurement. From experience and periodical verification data of the comparator, however, the effects due to Abbe errors are estimated to be within $\pm 0,02 \mu\text{m}$.

S13.9 Measurements (Δl)

The following observations are made of the inner diameter of the unknown and the setting ring:

No	Object	Observation	Measurand
1	reference setting ring	0 during this step the comparator display is zeroed	diameter in the nominal direction of the symmetry plane orthogonal to the cylinder axis
2	ring to be calibrated	49,999 35 mm	diameter in the nominal direction of the symmetry plane orthogonal to the cylinder axis
3	ring to be calibrated	49,999 11 mm	diameter in the symmetry plane orthogonal to the cylinder axis rotated around the axis with respect to the nominal direction by +1 mm on the circumference
4	ring to be calibrated	49,999 72 mm	diameter in the symmetry plane orthogonal to the cylinder axis rotated around the axis with respect to the nominal direction by -1 mm on the circumference.
5	ring to be calibrated	49,999 54 mm	diameter in the nominal direction translated to the plane parallel to symmetry plane orthogonal to the cylinder axis by 1 mm upwards
6	ring to be calibrated	49,999 96 mm	diameter in the nominal direction translated to the plane parallel to symmetry plane orthogonal to the cylinder axis by 1 mm downwards

The observations may be divided into two groups: the observation of the diameter of the setting ring (observation no 1) that is used to set the comparator display to zero and the observation of the diameter of the ring to be calibrated (observations no 2 to no 6) that give the difference in diameters:

arithmetic mean: $\bar{\Delta l} = 49,999\,54 \text{ mm}$

standard deviation of a single observation: $s(\Delta l) = 0,33 \mu\text{m}$

standard deviation of the mean: $s(\bar{\Delta l}) = \frac{s(\Delta l)}{\sqrt{5}} = 0,15 \mu\text{m}$

The standard deviation of a single observation $s(\Delta l) = 0,18 \mu\text{m}$ takes into account effects due to form deviations of the ring to be calibrated as well as due to the repeatability of the comparator. To obtain the standard uncertainty of measurement to be associated with the observed mean difference of the diameters, the uncertainty resulting from the zeroing of the comparator display must also be taken into account. This is deduced from the pooled estimate of the standard deviation $s_p(0) = 0,25 \mu\text{m}$ obtained in a prior measurement under the same conditions of measurement. The resulting standard measurement uncertainty to be associated with the observed diameter difference is:

$$u(\Delta l) = \sqrt{s^2(\bar{\Delta l}) + s_p^2(0)} = 0,30 \mu\text{m}$$

S13.10 Uncertainty budget (d_x)

quantity X_i	estimate x_i	standard uncertainty $u(x_i)$	probability distribution	sensitivity coefficient c_i	uncertainty contribution $u_i(y)$
d_S	40,000 7 mm	0,10 μm	normal	1,0	0,10 μm
Δl	49,999 55 mm	0,30 μm	normal	1,0	0,30 μm
δl_i	0,0 mm	0,22 μm	rectangular	1,0	0,22 μm
δl_T	0,0 mm	0,15 μm	normal	1,0	0,15 μm
δl_P	0,000 004 mm	0,0065 μm	rectangular	1,0	0,0065 μm
δl_E	0,0 mm	0,017 μm	rectangular	1,0	0,017 μm
δl_A	0,0 mm	0,012 μm	rectangular	1,0	0,012 μm
d_x	90,000 25 mm				0,411 μm

S13.11 Expanded uncertainty

$$U = k \cdot u(d_x) = 2 \times 0,411 \mu\text{m} \cong 0,9 \mu\text{m}$$

S13.12 Reported result

The diameter of the ring gauge is (90,000 3 \pm 0,000 9) mm.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

S13.13 Mathematical note on non-coaxiality

Since it is not possible to make an exact adjustment of the rings with respect to the measuring axis of the comparator, the quantity determined in the measurement is a chord of the respective ring in the proximity of its diameter. The length d' of this chord, which is observed in the measurement, is related to the diameter of the ring d by

$$d' = d \cdot \cos(\delta\varphi) \cong d \cdot \left(1 - \frac{1}{2}(\delta\varphi)^2\right) \quad (\text{S13.6})$$

where $\delta\varphi$ is the small angle that complements half of the central angle of the chord to $\pi/2$. This angle is related on the other hand to the small distance δc of the chord from the centre of the ring by

$$\delta c = \frac{1}{2} \cdot d \cdot \sin(\delta\varphi) \cong \frac{1}{2} \cdot d \cdot \delta\varphi \quad (\text{S13.7})$$

so that eq. (S13.6) may be rewritten as

$$d' \cong d - 2 \frac{(\delta c)^2}{D} \quad (\text{S13.8})$$

where the diameter d of the ring in the ratio has been replaced by its nominal diameter D since the nominator of the ratio is a small quantity already. The best estimate of the diameter is obtained by taking the expectation of the last relation to be

$$d = d' + 2 \frac{u^2(\delta c)}{D} . \quad (\text{S13.9})$$

Here it has been taken into account that the small distance δc has zero expectation. It must also be kept in mind that the meaning of d , d' and δc in eq. (S13.8) and eq. (S13.9) is not identical; whereas in eq. (S13.8) these symbols represent the not-exactly known quantities or random variables, in eq. (S13.9) they stand for the expectations of these quantities. Since the variance of a random variable equals the expectation of the square of its deviation from the respective expectation, the square of the standard measurement uncertainty to be associated with the diameter of the ring is, according to eq. (S13.8),

$$u^2(d) = u^2(d') + 4 \cdot (\alpha - 1) \frac{u^4(\delta c)}{D^2} \quad (\text{S13.10})$$

with

$$\alpha = \frac{m_4(\delta c)}{m_2^2(\delta c)} \quad (\text{S13.11})$$

being the ratio of the 4th order centred moment to the square of its 2nd order centred moment of the small distance δc . This ratio depends on the distribution that is assumed for δc . It takes the value $\alpha = 9/5$ if δc is assumed to be rectangularly distributed so that in this case the standard measurement uncertainty to be associated with the diameter is expressed by

$$u^2(d) = u^2(d') + \frac{16}{5} \cdot \frac{u^4(\delta c)}{D^2} \quad (\text{S13.12})$$